284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RTHR01

Lab Code: CHEM **Case No.:** P5093 **SAS No.:** P5093 **SDG NO.:** P5093

Continuing Calib Date: 12/06/2024 **Initial Calibration Date(s):** 11/08/2024 11/08/2024

Continuing Calib Time: 21:16 **Initial Calibration Time(s):** 02:45 10:39

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1242-1	(1)	5.93	5.92	5.82	6.02	-0.01
Aroclor-1242-2	(2)	5.95	5.95	5.85	6.05	0.00
Aroclor-1242-3	(3)	6.01	6.01	5.91	6.11	0.00
Aroclor-1242-4	(4)	6.11	6.11	6.01	6.21	0.00
Aroclor-1242-5	(5)	6.84	6.84	6.74	6.94	0.00
Tetrachloro-m-xylene		4.76	4.76	4.66	4.86	0.00
Decachlorobiphenyl		10.68	10.67	10.57	10.77	-0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RTHR01

Lab Code: CHEM **Case No.:** P5093 **SAS No.:** P5093 **SDG NO.:** P5093

Continuing Calib Date: 12/06/2024 **Initial Calibration Date(s):** 11/08/2024 11/08/2024

Continuing Calib Time: 21:16 **Initial Calibration Time(s):** 02:45 10:39

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1242-1	(1)	5.15	5.16	5.06	5.26	0.01
Aroclor-1242-2	(2)	5.17	5.18	5.08	5.28	0.01
Aroclor-1242-3	(3)	5.35	5.36	5.26	5.46	0.01
Aroclor-1242-4	(4)	5.44	5.44	5.34	5.54	0.00
Aroclor-1242-5	(5)	5.97	5.97	5.87	6.07	0.00
Tetrachloro-m-xylene		4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl		9.20	9.21	9.11	9.31	0.01