



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5093

Client: R3M Engineering, Inc.

Analytical Method: 8082A Datafile : PP068860.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165429BS	AR1016	5	4.30	ug/L	86				40 (61)	140 (112)	
	AR1260	5	4.10	ug/L	82				40 (66)	140 (113)	