

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: RTHR01

Lab Code: CHEM **Case No.:** P5093 **SAS No.:** P5093 **SDG NO.:** P5093

Instrument ID: ECD_D **Date(s) Analyzed:** 11/27/2024 11/27/2024

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.25	6.15	6.35	26714900
		2	6.45	6.35	6.55	37423900
		3	7.16	7.06	7.26	74221400
		4	7.57	7.47	7.67	97471900
		5	7.94	7.84	8.04	55327400

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: RTHR01

Lab Code: CHEM Case No.: P5093 SAS No.: P5093 SDG NO.: P5093

Instrument ID: ECD_D Date(s) Analyzed: 11/27/2024 11/27/2024

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.49	5.39	5.59	116683000
		2	5.66	5.56	5.76	78971400
		3	6.77	6.67	6.87	387418000
		4	7.21	7.11	7.31	277118000
		5	7.34	7.24	7.44	199715000