

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5093

Client: R3M Engineering, Inc.

Analytical Method: 8081B

Datafile : PD087060.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165431BSD	alpha-BHC	0.5	0.51	ug/L	101	1			40 (85)	140 (130)	20 (20)
	beta-BHC	0.5	0.51	ug/L	102	0			40 (83)	140 (126)	20 (20)
	delta-BHC	0.5	0.48	ug/L	96	1			40 (69)	140 (141)	20 (20)
	gamma-BHC (Lindane)	0.5	0.49	ug/L	99	1			40 (82)	140 (129)	20 (20)
	Heptachlor	0.5	0.50	ug/L	101	1			40 (79)	140 (127)	20 (20)
	Aldrin	0.5	0.48	ug/L	97	0			40 (79)	140 (126)	20 (20)
	Heptachlor epoxide	0.5	0.49	ug/L	97	1			40 (81)	140 (124)	20 (20)
	Endosulfan I	0.5	0.49	ug/L	99	1			40 (85)	140 (122)	20 (20)
	Dieldrin	0.5	0.50	ug/L	99	1			40 (83)	140 (125)	20 (20)
	4,4'-DDE	0.5	0.50	ug/L	99	1			40 (80)	140 (127)	20 (20)
	Endrin	0.5	0.48	ug/L	96	0			40 (81)	140 (128)	20 (20)
	Endosulfan II	0.5	0.50	ug/L	100	1			40 (82)	140 (123)	20 (20)
	4,4'-DDD	0.5	0.50	ug/L	100	1			40 (77)	140 (131)	20 (20)
	Endosulfan sulfate	0.5	0.48	ug/L	96	1			40 (76)	140 (129)	20 (20)
	4,4'-DDT	0.5	0.49	ug/L	98	1			40 (80)	140 (133)	20 (20)
	Methoxychlor	0.5	0.47	ug/L	94	0			40 (78)	140 (108)	20 (20)
	Endrin ketone	0.5	0.48	ug/L	96	1			40 (80)	140 (131)	20 (20)
	Endrin aldehyde	0.5	0.47	ug/L	94	1			40 (82)	140 (127)	20 (20)
	alpha-Chlordane	0.5	0.50	ug/L	99	1			40 (82)	140 (125)	20 (20)
	gamma-Chlordane	0.5	0.50	ug/L	101	1			40 (82)	140 (125)	20 (20)