

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120624\
 Data File : BF140760.D
 Acq On : 06 Dec 2024 12:08
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 06 11:51:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	88	0.00
2	1,4-Dioxane	40.000	39.896	0.3	89	-0.02
3	Pyridine	40.000	41.413	-3.5	83	-0.01
4	n-Nitrosodimethylamine	40.000	37.945	5.1	82	0.00
5 S	2-Fluorophenol	80.000	83.162	-4.0	92	0.00
6	Aniline	40.000	42.142	-5.4	81	0.00
7 S	Phenol-d6	80.000	82.586	-3.2	88	0.00
8	2-Chlorophenol	40.000	42.168	-5.4	92	0.00
9	Benzaldehyde	40.000	34.021	14.9	84	0.00
10 C	Phenol	40.000	41.507	-3.8	87	0.00
11	bis(2-Chloroethyl)ether	40.000	42.965	-7.4	92	0.00
12	1,3-Dichlorobenzene	40.000	42.028	-5.1	94	0.00
13 C	1,4-Dichlorobenzene	40.000	42.135	-5.3	94	0.00
14	1,2-Dichlorobenzene	40.000	42.351	-5.9	93	0.00
15	Benzyl Alcohol	40.000	40.164	-0.4	83	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.013	-2.5	89	0.00
17	2-Methylphenol	40.000	41.575	-3.9	89	0.00
18	Hexachloroethane	40.000	41.656	-4.1	92	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.650	-4.1	89	0.00
20	3+4-Methylphenols	40.000	42.961	-7.4	91	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	85	0.00
22	Acetophenone	40.000	42.552	-6.4	91	0.00
23 S	Nitrobenzene-d5	80.000	82.527	-3.2	87	0.00
24	Nitrobenzene	40.000	41.340	-3.4	88	0.00
25	Isophorone	40.000	42.182	-5.5	88	0.00
26 C	2-Nitrophenol	40.000	43.892	-9.7	92	0.00
27	2,4-Dimethylphenol	40.000	44.847	-12.1	91	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.386	-8.5	92	0.00
29 C	2,4-Dichlorophenol	40.000	43.683	-9.2	93	0.02
30	1,2,4-Trichlorobenzene	40.000	42.979	-7.4	93	0.00
31	Naphthalene	40.000	42.579	-6.4	92	0.00
32	Benzoic acid	40.000	44.312	-10.8	98	0.03
33	4-Chloroaniline	40.000	44.466	-11.2	89	0.01
34 C	Hexachlorobutadiene	40.000	43.124	-7.8	93	0.00
35	Caprolactam	40.000	42.490	-6.2	88	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.811	-7.0	88	0.00
37	2-Methylnaphthalene	40.000	43.448	-8.6	93	0.00
38	1-Methylnaphthalene	40.000	43.322	-8.3	92	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	85	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.527	-6.3	94	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.236	-5.6	95	0.00
42 S	2,4,6-Tribromophenol	80.000	86.678	-8.3	92	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.399	-6.0	91	0.00
44	2,4,5-Trichlorophenol	40.000	42.166	-5.4	89	0.02
45 S	2-Fluorobiphenyl	80.000	83.136	-3.9	92	0.00
46	1,1'-Biphenyl	40.000	42.115	-5.3	93	0.00
47	2-Chloronaphthalene	40.000	42.156	-5.4	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120624\
 Data File : BF140760.D
 Acq On : 06 Dec 2024 12:08
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 06 11:51:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.151	-0.4	85	0.00
49	Acenaphthylene	40.000	41.916	-4.8	92	0.00
50	Dimethylphthalate	40.000	42.765	-6.9	93	0.00
51	2,6-Dinitrotoluene	40.000	42.697	-6.7	91	0.00
52 C	Acenaphthene	40.000	41.551	-3.9	91	0.00
53	3-Nitroaniline	40.000	42.347	-5.9	88	0.01
54 P	2,4-Dinitrophenol	40.000	51.248	-28.1#	114	0.01
55	Dibenzofuran	40.000	40.992	-2.5	89	0.00
56 P	4-Nitrophenol	40.000	50.328	-25.8#	103	0.02
57	2,4-Dinitrotoluene	40.000	43.077	-7.7	90	0.00
58	Fluorene	40.000	42.119	-5.3	92	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	46.618	-16.5	98	0.00
60	Diethylphthalate	40.000	43.342	-8.4	94	0.00
61	4-Chlorophenyl-phenylether	40.000	44.094	-10.2	96	0.00
62	4-Nitroaniline	40.000	41.509	-3.8	87	0.01
63	Azobenzene	40.000	42.261	-5.7	93	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	87	0.00
65	4,6-Dinitro-2-methylphenol	40.000	48.729	-21.8	98	0.01
66 c	n-Nitrosodiphenylamine	40.000	42.074	-5.2	93	0.00
67	4-Bromophenyl-phenylether	40.000	44.407	-11.0	99	0.00
68	Hexachlorobenzene	40.000	42.749	-6.9	95	0.00
69	Atrazine	40.000	33.976	15.1	87	0.00
70 C	Pentachlorophenol	40.000	41.011	-2.5	80	0.00
71	Phenanthrene	40.000	42.080	-5.2	93	0.00
72	Anthracene	40.000	42.585	-6.5	94	0.00
73	Carbazole	40.000	42.809	-7.0	94	0.00
74	Di-n-butylphthalate	40.000	46.268	-15.7	102	0.00
75 C	Fluoranthene	40.000	44.016	-10.0	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	95	0.00
77	Benzydine	40.000	38.204	4.5	92	0.01
78	Pyrene	40.000	41.551	-3.9	98	0.00
79 S	Terphenyl-d14	80.000	84.481	-5.6	100	0.00
80	Butylbenzylphthalate	40.000	44.480	-11.2	103	0.00
81	Benzo(a)anthracene	40.000	42.005	-5.0	102	0.00
82	3,3'-Dichlorobenzidine	40.000	42.973	-7.4	98	0.00
83	Chrysene	40.000	42.841	-7.1	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.886	-7.2	103	0.00
85 c	Di-n-octyl phthalate	40.000	44.010	-10.0	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	92	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	43.474	-8.7	100	0.04
88	Benzo(b)fluoranthene	40.000	42.513	-6.3	104	0.02
89	Benzo(k)fluoranthene	40.000	43.392	-8.5	100	0.02
90 C	Benzo(a)pyrene	40.000	42.833	-7.1	100	0.02
91	Dibenzo(a,h)anthracene	40.000	44.209	-10.5	102	0.04
92	Benzo(g,h,i)perylene	40.000	43.853	-9.6	101	0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120624\
Data File : BF140760.D
Acq On : 06 Dec 2024 12:08
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 06 11:51:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 21 15:23:48 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0