

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112124\
 Data File : BF140533.D
 Acq On : 21 Nov 2024 13:25
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Quant Time: Nov 21 15:13:26 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 14:50:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	117539	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	425301	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	233869	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	432727	20.000	ng	0.00
76) Chrysene-d12	14.057	240	241266	20.000	ng	0.00
86) Perylene-d12	15.545	264	219139	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	656695	95.327	ng	0.00
7) Phenol-d6	6.522	99	860924	94.531	ng	0.00
23) Nitrobenzene-d5	7.445	82	801144	96.352	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	245159	98.015	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	1467020	93.462	ng	0.00
79) Terphenyl-d14	12.992	244	1481875	95.640	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.646	88	132951	45.902	ng	99
3) Pyridine	3.416	79	312094	48.973	ng	97
4) n-Nitrosodimethylamine	3.387	42	185759	49.595	ng	99
6) Aniline	6.545	93	306198	47.767	ng	99
8) 2-Chlorophenol	6.669	128	352950	47.623	ng	97
9) Benzaldehyde	6.428	77	220959	45.830	ng	100
10) Phenol	6.534	94	444859	47.830	ng	98
11) bis(2-Chloroethyl)ether	6.616	93	336892	47.335	ng	100
12) 1,3-Dichlorobenzene	6.816	146	395334	47.472	ng	99
13) 1,4-Dichlorobenzene	6.892	146	399191	47.341	ng	100
14) 1,2-Dichlorobenzene	7.045	146	372552	47.151	ng	98
15) Benzyl Alcohol	7.022	79	327081	48.428	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.151	45	392372	46.665	ng	93
17) 2-Methylphenol	7.134	107	280882	47.344	ng	98
18) Hexachloroethane	7.386	117	151045	47.961	ng	99
19) n-Nitroso-di-n-propyla...	7.292	70	251453	46.718	ng	99
20) 3+4-Methylphenols	7.292	107	355860	46.666	ng	95
22) Acetophenone	7.286	105	492194	47.469	ng	98
24) Nitrobenzene	7.463	77	412860	48.043	ng	99
25) Isophorone	7.698	82	668649	48.214	ng	100
26) 2-Nitrophenol	7.775	139	189056	49.644	ng	99
27) 2,4-Dimethylphenol	7.816	122	217017	47.628	ng	99
28) bis(2-Chloroethoxy)met...	7.904	93	402409	47.630	ng	100
29) 2,4-Dichlorophenol	8.022	162	295023	48.863	ng	98
30) 1,2,4-Trichlorobenzene	8.098	180	336812	48.858	ng	98
31) Naphthalene	8.181	128	1055423	48.178	ng	100
32) Benzoic acid	7.957	122	196545	49.333	ng	99
33) 4-Chloroaniline	8.233	127	322212	49.125	ng	99
34) Hexachlorobutadiene	8.292	225	221707	48.555	ng	99
35) Caprolactam	8.616	113	90596	48.486	ng	99
36) 4-Chloro-3-methylphenol	8.716	107	326418	48.287	ng	99
37) 2-Methylnaphthalene	8.869	142	662592	47.620	ng	100
38) 1-Methylnaphthalene	8.969	142	649248	47.603	ng	99
40) 1,2,4,5-Tetrachloroben...	9.033	216	328190	47.935	ng	99
41) Hexachlorocyclopentadiene	9.016	237	70831	49.123	ng	97
43) 2,4,6-Trichlorophenol	9.151	196	212974	49.574	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112124\
 Data File : BF140533.D
 Acq On : 21 Nov 2024 13:25
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Quant Time: Nov 21 15:13:26 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 14:50:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	227840	48.866	ng	98
46) 1,1'-Biphenyl	9.333	154	834139	47.772	ng	99
47) 2-Chloronaphthalene	9.363	162	632459	47.803	ng	99
48) 2-Nitroaniline	9.457	65	212420	50.020	ng	98
49) Acenaphthylene	9.775	152	951784	47.612	ng	99
50) Dimethylphthalate	9.639	163	740559	48.081	ng	99
51) 2,6-Dinitrotoluene	9.704	165	169989	48.663	ng	96
52) Acenaphthene	9.951	154	615149	47.524	ng	99
53) 3-Nitroaniline	9.875	138	169234	49.534	ng	96
54) 2,4-Dinitrophenol	9.980	184	84985	50.205	ng	99
55) Dibenzofuran	10.122	168	911771	47.061	ng	99
56) 4-Nitrophenol	10.045	139	124021	53.227	ng	98
57) 2,4-Dinitrotoluene	10.110	165	225762	48.629	ng	98
58) Fluorene	10.463	166	738453	47.486	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.245	232	178549	49.828	ng	97
60) Diethylphthalate	10.333	149	755358	48.268	ng	100
61) 4-Chlorophenyl-phenyle...	10.451	204	361856	47.414	ng	100
62) 4-Nitroaniline	10.492	138	180999	50.512	ng	99
63) Azobenzene	10.616	77	704370	47.529	ng	99
65) 4,6-Dinitro-2-methylph...	10.522	198	119290	53.947	ng	99
66) n-Nitrosodiphenylamine	10.574	169	623715	48.740	ng	100
67) 4-Bromophenyl-phenylether	10.945	248	216523	48.587	ng	97
68) Hexachlorobenzene	11.016	284	251498	48.739	ng	100
69) Atrazine	11.098	200	159048	44.180	ng	99
70) Pentachlorophenol	11.210	266	124128	54.656	ng	99
71) Phenanthrene	11.427	178	1000897	48.118	ng	100
72) Anthracene	11.480	178	978923	48.111	ng	100
73) Carbazole	11.639	167	948090	48.436	ng	99
74) Di-n-butylphthalate	11.957	149	1105032	48.899	ng	100
75) Fluoranthene	12.621	202	1080641	47.849	ng	100
77) Benzidine	12.745	184	442856	63.133	ng	100
78) Pyrene	12.851	202	1080149	48.420	ng	100
80) Butylbenzylphthalate	13.463	149	394867	49.136	ng	99
81) Benzo(a)anthracene	14.045	228	780853	48.892	ng	100
82) 3,3'-Dichlorobenzidine	14.004	252	244892	51.072	ng	99
83) Chrysene	14.080	228	685847	46.964	ng	100
84) Bis(2-ethylhexyl)phtha...	14.021	149	497161	48.994	ng	100
85) Di-n-octyl phthalate	14.645	149	692020	49.901	ng	100
87) Indeno(1,2,3-cd)pyrene	17.062	276	715720	50.117	ng	100
88) Benzo(b)fluoranthene	15.109	252	683919	49.687	ng	100
89) Benzo(k)fluoranthene	15.139	252	559704	46.467	ng	100
90) Benzo(a)pyrene	15.486	252	546539	48.835	ng	99
91) Dibenzo(a,h)anthracene	17.074	278	588742	50.339	ng	99
92) Benzo(g,h,i)perylene	17.521	276	599158	50.203	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 21 15:13:26 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 21 14:50:10 2024
Response via : Initial Calibration

