

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112124

Review By	yogesh	Review On	11/22/2024 3:53:36 AM
Supervise By	mohammad	Supervise On	11/27/2024 6:04:00 AM
SubDirectory	BF112124	HP Acquire Method	BNA_F
		HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12329,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140526.D	21 Nov 2024 10:17		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140527.D	21 Nov 2024 10:44	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BF140528.D	21 Nov 2024 11:13		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BF140529.D	21 Nov 2024 11:39	Compound#9,32,41,54,56,65,70 removed from 5 ppm	RC/JU	Ok
5	SSTDICC010	SSTDICC010	BF140530.D	21 Nov 2024 12:05		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BF140531.D	21 Nov 2024 12:32	Compound#32,41,54 Kept on LR	RC/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BF140532.D	21 Nov 2024 12:58	Calibration failed for Benzidine	RC/JU	Ok,M
8	SSTDICC050	SSTDICC050	BF140533.D	21 Nov 2024 13:25	The Calibration is Good For 8270 DOD Except com#77 and good for 625.1 Method Except com#77	RC/JU	Ok
9	SSTDICC060	SSTDICC060	BF140534.D	21 Nov 2024 13:51		RC/JU	Ok
10	SSTDICC080	SSTDICC080	BF140535.D	21 Nov 2024 14:18	Compound#9 removed from 80 ppm	RC/JU	Ok
11	SSTDICV040	ICVBF112124	BF140536.D	21 Nov 2024 15:07		RC/JU	Ok,M
12	PB165144BL	PB165144BL	BF140537.D	21 Nov 2024 15:34		RC/JU	Ok
13	DFTPP	DFTPP	BF140538.D	21 Nov 2024 16:27		RC/JU	Ok
14	SSTDCCC040	SSTDCCC040	BF140539.D	21 Nov 2024 16:54		RC/JU	Ok,M
15	PB165060TB	PB165060TB	BF140540.D	21 Nov 2024 17:20		RC/JU	Ok
16	P4887-06	MH-760	BF140541.D	21 Nov 2024 17:55		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112124

Review By	yogesh	Review On	11/22/2024 3:53:36 AM
Supervise By	mohammad	Supervise On	11/27/2024 6:04:00 AM
SubDirectory	BF112124	HP Acquire Method	BNA_F
		HP Processing Method	bf112124
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621		
CCC	SP6624		
Internal Standard/PEM	S12329,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

17	P4887-02	MH-739	BF140542.D	21 Nov 2024 18:21		RC/JU	Ok
18	P4870-16	TP-15	BF140543.D	21 Nov 2024 18:48		RC/JU	Ok
19	P4870-15	MH-736	BF140544.D	21 Nov 2024 19:14		RC/JU	Ok
20	P4870-14	MH-735	BF140545.D	21 Nov 2024 19:40		RC/JU	Ok
21	P4870-13	TP-1	BF140546.D	21 Nov 2024 20:07		RC/JU	Ok
22	P4929-02	ARS520	BF140547.D	21 Nov 2024 20:33		RC/JU	Ok
23	P4924-04	MH-4	BF140548.D	21 Nov 2024 20:59		RC/JU	Ok
24	P4921-01	WC-11-A-202411	BF140549.D	21 Nov 2024 21:25		RC/JU	Ok
25	P4916-12	TP-3-WC	BF140550.D	21 Nov 2024 21:51		RC/JU	Ok
26	P4916-08	TP-2-WC	BF140551.D	21 Nov 2024 22:18		RC/JU	Ok
27	P4916-04	TP-1-WC	BF140552.D	21 Nov 2024 22:44		RC/JU	Ok
28	P4887-01	MH-739	BF140553.D	21 Nov 2024 23:10		RC/JU	Ok
29	P4916-09	TP-3-WC	BF140554.D	21 Nov 2024 23:36		RC/JU	Ok
30	P4916-09MS	TP-3-WCMS	BF140555.D	22 Nov 2024 00:02		RC/JU	Ok,M
31	P4916-09MSD	TP-3-WCMSD	BF140556.D	22 Nov 2024 00:28		RC/JU	Ok,M
32	P4916-01	TP-1-WC	BF140557.D	22 Nov 2024 00:54		RC/JU	Ok
33	P4916-05RE	TP-2-WCRE	BF140558.D	22 Nov 2024 01:21	Internal Standard Fail	RC/JU	Confirms
34	P4887-05RE	MH-760RE	BF140559.D	22 Nov 2024 01:47	Internal Standard Fail	RC/JU	Confirms
35	P4924-01	MH-4	BF140560.D	22 Nov 2024 02:13	Internal Standard Fail	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112124

Review By	yogesh	Review On	11/22/2024 3:53:36 AM
Supervise By	mohammad	Supervise On	11/27/2024 6:04:00 AM
SubDirectory	BF112124	HP Acquire Method	BNA_F
		HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12329,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

36	P4936-01	PL-01-11202024	BF140561.D	22 Nov 2024 02:39	Internal Standard Fail	RC/JU	Ok,M
37	P4929-01	ARS520	BF140562.D	22 Nov 2024 03:06	Internal Standard Fail	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF120624

Review By	yogesh	Review On	12/10/2024 12:52:35 AM
Supervise By	mohammad	Supervise On	12/10/2024 1:02:52 AM
SubDirectory	BF120624	HP Acquire Method	BNA_F
		HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12331,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140759.D	06 Dec 2024 11:41		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140760.D	06 Dec 2024 12:08		RC/JU	Ok,M
3	PB165423BL	PB165423BL	BF140761.D	06 Dec 2024 12:51		RC/JU	Ok
4	PB165423BS	PB165423BS	BF140762.D	06 Dec 2024 13:16		RC/JU	Ok,M
5	P5133-01	MOO-24-00374	BF140763.D	06 Dec 2024 13:48	Internal Standard Fail	RC/JU	Ok
6	P5137-01	LAW-OILY-STONES	BF140764.D	06 Dec 2024 14:14	Internal Standard Fail	RC/JU	Ok
7	P5093-01	LL-001	BF140765.D	06 Dec 2024 14:40		RC/JU	Ok
8	P5093-02	LL-001-FB-12-4-24	BF140766.D	06 Dec 2024 15:06		RC/JU	Ok
9	P5145-01	286085	BF140767.D	06 Dec 2024 15:32	Internal Standard Fail	RC/JU	Ok
10	P5095-04	MH-764	BF140768.D	06 Dec 2024 15:58		RC/JU	Ok
11	P5095-04MS	MH-764MS	BF140769.D	06 Dec 2024 16:24	END CCC Missing	RC/JU	Not Ok
12	P5095-04MSD	MH-764MSD	BF140770.D	06 Dec 2024 16:50	END CCC Missing	RC/JU	Not Ok
13	P5096-04	MH-B	BF140771.D	06 Dec 2024 17:16		RC/JU	Ok
14	P5096-08	MH-A	BF140772.D	06 Dec 2024 17:42		RC/JU	Ok
15	P5117-02	TAPIAL2-IDW-SOIL-12	BF140773.D	06 Dec 2024 18:08	END CCC Missing	RC/JU	Not Ok
16	P5133-02	MOO-24-00374	BF140774.D	06 Dec 2024 18:34		RC/JU	Ok
17	P5136-02	COMP-1	BF140775.D	06 Dec 2024 19:00		RC/JU	Ok
18	P5096-01DL	MH-BDL	BF140776.D	06 Dec 2024 19:26		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF120624

Review By	yogesh	Review On	12/10/2024 12:52:35 AM		
Supervise By	mohammad	Supervise On	12/10/2024 1:02:52 AM		
SubDirectory	BF120624	HP Acquire Method	BNA_F	HP Processing Method	bf112124
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12331,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P5108-01	ASPHALT-COMP	BF140777.D	06 Dec 2024 19:52		RC/JU	Ok
20	P5136-01	COMP-1	BF140778.D	06 Dec 2024 20:18	Internal Standard Fail	RC/JU	Ok
21	P5112-01	10TH-ST-SOIL	BF140779.D	06 Dec 2024 20:44	Internal Standard Fail, Need 5X Dilution	RC/JU	Dilution
22	P5112-01MS	10TH-ST-SOILMS	BF140780.D	06 Dec 2024 21:10		RC/JU	Ok,M
23	P5112-01MSD	10TH-ST-SOILMSD	BF140781.D	06 Dec 2024 21:36		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF120924

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF120924	HP Acquire Method	BNA_F	HP Processing Method	bf112124
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12331,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140782.D	09 Dec 2024 11:55		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140783.D	09 Dec 2024 12:21		RC/JU	Ok,NR
3	PB165432BL	PB165432BL	BF140784.D	09 Dec 2024 12:47		RC/JU	Ok
4	PB165432BS	PB165432BS	BF140785.D	09 Dec 2024 13:13		RC/JU	Ok,NR
5	PB165432BSD	PB165432BSD	BF140786.D	09 Dec 2024 13:39		RC/JU	Ok,NR
6	PB165435BL	PB165435BL	BF140787.D	09 Dec 2024 14:05		RC/JU	Ok
7	PB165435BS	PB165435BS	BF140788.D	09 Dec 2024 14:31		RC/JU	Ok,NR
8	PB165390TB	PB165390TB	BF140789.D	09 Dec 2024 14:58		RC/JU	Ok
9	P5117-02	TAPIAL2-IDW-SOIL-12	BF140790.D	09 Dec 2024 15:29		RC/JU	Ok
10	P5095-04MS	MH-764MS	BF140791.D	09 Dec 2024 15:55		RC/JU	Ok,NR
11	P5095-04MSD	MH-764MSD	BF140792.D	09 Dec 2024 16:22		RC/JU	Ok,NR
12	SSTDCCC040	SSTDCCC040EC	BF140793.D	09 Dec 2024 16:54		RC/JU	Ok,NR

M : Manual Integration