

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5093 SAS No.: P5093 SDG NO.: P5093

EPA Sample No.: SSTDCCC040 Date Analyzed: 12/09/2024

Lab File ID: BF140783.D Time Analyzed: 12:21

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	92885	6.863	346239	8.15	189814	9.90
UPPER LIMIT	185770	7.363	692478	8.645	379628	10.404
LOWER LIMIT	46442.5	6.363	173120	7.645	94907	9.404
EPA SAMPLE NO.						
01 PB165432BL	69444	6.86	265288	8.15	149935	9.90
02 PB165432BS	74041	6.86	285606	8.15	160534	9.90
03 PB165432BSD	73193	6.86	278673	8.15	158385	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.