

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5093 SAS No.: P5093 SDG NO.: P5093

EPA Sample No.: SSTDCCC040 Date Analyzed: 12/06/2024

Lab File ID: BF140760.D Time Analyzed: 12:08

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	378576	11.404	226669	14.057	194773	15.557
UPPER LIMIT	757152	11.904	453338	14.557	389546	16.057
LOWER LIMIT	189288	10.904	113335	13.557	97386.5	15.057
EPA SAMPLE NO.						
01 LL-001	232952	11.40	138014	14.05	150336	15.56
02 LL-001-FB-12-4-24	294061	11.40	155103	14.05	186876	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.