

Surrogate Summary

SDG No.: **P5093**

Client: **R3M Engineering, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5093-01	LL-001	1,2-Dichloroethane-d4	50	52.6	105	70 (74)	130 (125)
		Dibromofluoromethane	50	46.0	92	70 (75)	130 (124)
		Toluene-d8	50	50.2	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.3	99	70 (77)	130 (121)
P5093-02	LL-001-FB-12-4-24	1,2-Dichloroethane-d4	50	53.4	107	70 (74)	130 (125)
		Dibromofluoromethane	50	45.7	91	70 (75)	130 (124)
		Toluene-d8	50	48.7	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.4	99	70 (77)	130 (121)
P5093-03	TB-12-4-24	1,2-Dichloroethane-d4	50	53.0	106	70 (74)	130 (125)
		Dibromofluoromethane	50	45.3	91	70 (75)	130 (124)
		Toluene-d8	50	48.9	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.3	99	70 (77)	130 (121)
VX1204WBL01	VX1204WBL01	1,2-Dichloroethane-d4	50	52.1	104	70 (74)	130 (125)
		Dibromofluoromethane	50	46.1	92	70 (75)	130 (124)
		Toluene-d8	50	49.2	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.7	97	70 (77)	130 (121)
VX1204WBS01	VX1204WBS01	1,2-Dichloroethane-d4	50	56.1	112	70 (74)	130 (125)
		Dibromofluoromethane	50	51.6	103	70 (75)	130 (124)
		Toluene-d8	50	50.5	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.5	105	70 (77)	130 (121)
VX1204WBSD01	VX1204WBSD01	1,2-Dichloroethane-d4	50	55.3	111	70 (74)	130 (125)
		Dibromofluoromethane	50	52.9	106	70 (75)	130 (124)
		Toluene-d8	50	51.9	104	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.7	109	70 (77)	130 (121)

