

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RTHR01
 Lab Code: CHEM Case No.: P5093 SAS No.: P5093 SDG NO.: P5093
 Lab File ID: VX044109.D Date Analyzed: 12/04/2024
 Instrument ID: MSVOA_X Time Analyzed: 12:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	121433	5.54	212858	6.76	180347	10.05
UPPER LIMIT	242866	6.038	425716	7.257	360694	10.549
LOWER LIMIT	60716.5	5.038	106429	6.257	90173.5	9.549
EPA SAMPLE NO.						
LL-001	106401	5.54	205400	6.76	179121	10.06
LL-001-FB-12-4-24	110604	5.54	218721	6.76	189224	10.06
TB-12-4-24	110240	5.54	214620	6.76	184877	10.06
VX1204WBL01	113722	5.55	219894	6.76	188758	10.05
VX1204WBS01	120163	5.54	225592	6.76	189832	10.06
VX1204WBSD01	115232	5.54	206872	6.76	178913	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.