

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RTHR01
 Lab Code: CHEM Case No.: P5093 SAS No.: P5093 SDG NO.: P5093
 Lab File ID: VX044109.D Date Analyzed: 12/04/2024
 Instrument ID: MSVOA_X Time Analyzed: 12:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	85782	12.018				
UPPER LIMIT	171564	12.518				
LOWER LIMIT	42891	11.518				
EPA SAMPLE NO.						
LL-001	75267	12.02				
LL-001-FB-12-4-24	79833	12.02				
TB-12-4-24	78562	12.02				
VX1204WBL01	77704	12.02				
VX1204WBS01	84893	12.02				
VX1204WBSD01	82606	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.