

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RTHR01
Lab Code: CHEM Case No.: P5093 SAS No.: P5093 SDG NO.: P5093
Lab File ID: VX043924.D BFB Injection Date: 11/21/2024
Instrument ID: MSVOA_X BFB Injection Time: 08:51
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.4
75	30.0 - 60.0% of mass 95	54.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	70.5
175	5.0 - 9.0% of mass 174	4.8 (6.8) 1
176	95.0 - 101.0% of mass 174	67.9 (96.4) 1
177	5.0 - 9.0% of mass 176	4.9 (7.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX043925.D	11/21/2024	09:47
VSTDIC005	VSTDIC005	VX043926.D	11/21/2024	10:14
VSTDIC020	VSTDIC020	VX043927.D	11/21/2024	10:37
VSTDIC050	VSTDIC050	VX043928.D	11/21/2024	11:17
VSTDIC100	VSTDIC100	VX043929.D	11/21/2024	11:40
VSTDIC150	VSTDIC150	VX043930.D	11/21/2024	12:03