

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB133777

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/06/2024	08:55
2	CAL2	1	Water	NA	NA	20.2	7.00	12/06/2024	08:56
3	CAL3	1	Water	NA	NA	20.3	10.02	12/06/2024	08:58
4	ICV	1	Water	NA	NA	20.3	7.00	12/06/2024	09:00
5	CCV1	1	Water	NA	NA	20.1	2.01	12/06/2024	09:47
6	P5112-02	1	Solid	20.02	20	20.1	6.41	12/06/2024	09:55
7	P5112-02DUP	1	Solid	20.03	20	20.2	6.42	12/06/2024	09:56
8	P5133-02	1	Solid	20.03	20	20.4	6.07	12/06/2024	10:00
9	P5136-02	1	Solid	20.04	20	20.1	6.29	12/06/2024	10:05
10	CCV2	1	Water	NA	NA	20.3	12.02	12/06/2024	10:10

WORKLIST(Hardcopy Internal Chain)

VB 133777

WorkList Name : corrosivity p5133

WorkList ID : 186057

Department : Wet-Chemistry

Date : 12-06-2024 08:46:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5112-02	10TH-ST-SOIL	Solid	Corrosivity	Cool 4 deg C	TULL02	L51	12/05/2024	9045D
P5133-02	MOO-24-00374	Solid	Corrosivity	Cool 4 deg C	PSEG03	L61	12/05/2024	9045D
P5136-02	COMP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L61	12/05/2024	9045D

Date/Time 12/06/24 08:55
Raw Sample Received by: rd wllc
Raw Sample Relinquished by: [signature]

Date/Time 12/06/24 14:00
Raw Sample Received by: [signature]
Raw Sample Relinquished by: rd wllc