



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Tully Construction Co., Inc.

SDG No.: P5112

Project: 10th Street & 2nd Avenue

RunNo.: LB133773

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.097	0.099	98	85-115	12/05/2024
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	12/05/2024
Sample ID: CCV2 Reactive Cyanide	mg/L	0.23	0.25	92	90-110	12/05/2024
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	12/05/2024

Initial and Continuing Calibration Verification

Client: Tully Construction Co., Inc.

SDG No.: P5112

Project: 10th Street & 2nd Avenue

RunNo.: LB133777

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Corrosivity		pH	7.00	7	100	90-110	12/06/2024
Sample ID:	CCV1						
Corrosivity		pH	2.01	2.00	101	90-110	12/06/2024
Sample ID:	CCV2						
Corrosivity		pH	12.02	12.00	100	90-110	12/06/2024