



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Construction Co., Inc. **SDG No.:** P5112
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5112 **SAS No.:** P5112
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	102	40 - 160	CV	12/06/2024	16:56	LB133794
CRI01	Aluminum	105	100	104	40 - 160	P	12/09/2024	11:59	LB133834
	Antimony	50.2	50.0	100	40 - 160	P	12/09/2024	11:59	LB133834
	Arsenic	20.0	20.0	100	40 - 160	P	12/09/2024	11:59	LB133834
	Barium	96.4	100	96	40 - 160	P	12/09/2024	11:59	LB133834
	Beryllium	6.07	6.0	101	40 - 160	P	12/09/2024	11:59	LB133834
	Cadmium	5.65	6.0	94	40 - 160	P	12/09/2024	11:59	LB133834
	Calcium	2060	2000	103	40 - 160	P	12/09/2024	11:59	LB133834
	Chromium	9.89	10.0	99	40 - 160	P	12/09/2024	11:59	LB133834
	Cobalt	28.6	30.0	95	40 - 160	P	12/09/2024	11:59	LB133834
	Copper	21.8	20.0	109	40 - 160	P	12/09/2024	11:59	LB133834
	Iron	93.6	100	94	40 - 160	P	12/09/2024	11:59	LB133834
	Lead	11.0	12.0	91	40 - 160	P	12/09/2024	11:59	LB133834
	Magnesium	2160	2000	108	40 - 160	P	12/09/2024	11:59	LB133834
	Manganese	20.5	20.0	103	40 - 160	P	12/09/2024	11:59	LB133834
	Nickel	38.6	40.0	96	40 - 160	P	12/09/2024	11:59	LB133834
	Potassium	1910	2000	96	40 - 160	P	12/09/2024	11:59	LB133834
	Selenium	18.4	20.0	92	40 - 160	P	12/09/2024	11:59	LB133834
	Silver	9.73	10.0	97	40 - 160	P	12/09/2024	11:59	LB133834
	Sodium	1800	2000	90	40 - 160	P	12/09/2024	11:59	LB133834
	Thallium	40.0	40.0	100	40 - 160	P	12/09/2024	11:59	LB133834
	Vanadium	40.0	40.0	100	40 - 160	P	12/09/2024	11:59	LB133834
	Zinc	42.2	40.0	106	40 - 160	P	12/09/2024	11:59	LB133834