

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Construction Co., Inc. **SDG No.:** P5112
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5112 **SAS No.:** P5112
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	256000	255000	100	216000	294000	12/09/2024	12:08	LB133834
	Antimony	-0.86			-50	50	12/09/2024	12:08	LB133834
	Arsenic	3.35			-20	20	12/09/2024	12:08	LB133834
	Barium	5.62	6.0	94	-94	106	12/09/2024	12:08	LB133834
	Beryllium	1.49			-6	6	12/09/2024	12:08	LB133834
	Cadmium	3.33	1.0	333	-5	7	12/09/2024	12:08	LB133834
	Calcium	243000	245000	99	208000	282000	12/09/2024	12:08	LB133834
	Chromium	56.5	52.0	109	42	62	12/09/2024	12:08	LB133834
	Cobalt	2.04			-30	30	12/09/2024	12:08	LB133834
	Copper	18.7	2.0	935	-18	22	12/09/2024	12:08	LB133834
	Iron	95800	101000	95	85600	116500	12/09/2024	12:08	LB133834
	Lead	11.9			-12	12	12/09/2024	12:08	LB133834
	Magnesium	265000	255000	104	216000	294000	12/09/2024	12:08	LB133834
	Manganese	4.34	7.0	62	-13	27	12/09/2024	12:08	LB133834
	Nickel	31.5	2.0	1575	-38	42	12/09/2024	12:08	LB133834
	Potassium	18.3			0	0	12/09/2024	12:08	LB133834
	Selenium	-19.6			-20	20	12/09/2024	12:08	LB133834
	Silver	-4.87			-10	10	12/09/2024	12:08	LB133834
	Sodium	54.5			0	0	12/09/2024	12:08	LB133834
	Thallium	-3.39			-40	40	12/09/2024	12:08	LB133834
	Vanadium	9.50			-40	40	12/09/2024	12:08	LB133834
	Zinc	31.5			-40	40	12/09/2024	12:08	LB133834
ICSAB01	Aluminum	253000	247000	102	209000	285000	12/09/2024	12:17	LB133834
	Antimony	622	618	101	525	711	12/09/2024	12:17	LB133834
	Arsenic	105	104	101	88.4	120	12/09/2024	12:17	LB133834
	Barium	508	537	95	437	637	12/09/2024	12:17	LB133834
	Beryllium	522	495	106	420	570	12/09/2024	12:17	LB133834
	Cadmium	1000	972	103	826	1120	12/09/2024	12:17	LB133834
	Calcium	239000	235000	102	199000	271000	12/09/2024	12:17	LB133834
	Chromium	563	542	104	460	624	12/09/2024	12:17	LB133834
	Cobalt	507	476	106	404	548	12/09/2024	12:17	LB133834
	Copper	498	511	98	434	588	12/09/2024	12:17	LB133834
	Iron	95800	99300	96	84400	114500	12/09/2024	12:17	LB133834
	Lead	52.4	49.0	107	37	61	12/09/2024	12:17	LB133834
	Magnesium	259000	248000	104	210000	286000	12/09/2024	12:17	LB133834
	Manganese	494	507	97	430	584	12/09/2024	12:17	LB133834
	Nickel	1020	954	107	810	1100	12/09/2024	12:17	LB133834
	Potassium	-36.7			0	0	12/09/2024	12:17	LB133834
	Selenium	29.8	46.0	65	26	66	12/09/2024	12:17	LB133834
	Silver	195	201	97	170	232	12/09/2024	12:17	LB133834
	Sodium	5.96			0	0	12/09/2024	12:17	LB133834
	Thallium	81.8	108	76	68	148	12/09/2024	12:17	LB133834

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Construction Co., Inc. **SDG No.:** P5112
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5112 **SAS No.:** P5112
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	500	491	102	417	565	12/09/2024	12:17	LB133834
	Zinc	922	952	97	809	1095	12/09/2024	12:17	LB133834