

**Metals**  
**- 5b -**  
**POST DIGEST SPIKE SUMMARY**

**Client:** Tully Construction Co., Inc. **SDG No.:** P5112  
**Contract:** TULL02 **Lab Code:** CHEM **Case No.:** P5112 **SAS No.:** P5112  
**Matrix:** Solid **Level:** LOW **Client ID:** 10TH-ST-SOILA  
**Sample ID:** P5112-01 **Spiked ID:** P5112-01A

| Analyte  | Units | Acceptance<br>Limit %R | Spiked<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|----------|-------|------------------------|------------------|---|------------------|---|----------------|---------------|------|---|
| Chromium | mg/Kg | 75 - 125               | 38.6             |   | 25.1             |   | 19.4           | 70            |      | P |
| Copper   | mg/Kg | 75 - 125               | 55.4             |   | 36.4             |   | 14.5           | 131           |      | P |
| Sodium   | mg/Kg | 75 - 125               | 217              |   | 109              |   | 150            | 72            |      | P |