

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Construction Co., Inc. **SDG No.:** P5112
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5112 **SAS No.:** P5112

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165416BS							
Aluminum	mg/Kg	80.6	77.9		97	80 - 120	P
Antimony	mg/Kg	32.3	31.2		97	80 - 120	P
Arsenic	mg/Kg	32.3	29.9		93	80 - 120	P
Barium	mg/Kg	8.1	7.75		96	80 - 120	P
Beryllium	mg/Kg	8.1	7.76		96	80 - 120	P
Cadmium	mg/Kg	8.1	7.45		92	80 - 120	P
Calcium	mg/Kg	40.3	38.6	J	96	80 - 120	P
Chromium	mg/Kg	16.1	15.3		95	80 - 120	P
Cobalt	mg/Kg	8.1	7.55		93	80 - 120	P
Copper	mg/Kg	12.1	12.2		101	80 - 120	P
Iron	mg/Kg	120	112		93	80 - 120	P
Lead	mg/Kg	40.3	36.9		92	80 - 120	P
Magnesium	mg/Kg	80.6	74.5	J	92	80 - 120	P
Manganese	mg/Kg	8.1	7.82		96	80 - 120	P
Nickel	mg/Kg	20.2	18.9		94	80 - 120	P
Potassium	mg/Kg	400	371		93	80 - 120	P
Selenium	mg/Kg	80.6	76.4		95	80 - 120	P
Silver	mg/Kg	3.0	2.85		95	80 - 120	P
Sodium	mg/Kg	120	110		92	80 - 120	P
Thallium	mg/Kg	80.6	78.8		98	80 - 120	P
Vanadium	mg/Kg	12.1	11.6		96	80 - 120	P
Zinc	mg/Kg	8.1	8.02		99	80 - 120	P



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Construction Co., Inc. SDG No.: P5112
Contract: TULL02 Lab Code: CHEM Case No.: P5112 SAS No.: P5112

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165448BS Mercury	mg/Kg	0.27	0.29		106	80 - 124	CV