

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Tully Construction Co., Inc. **SDG No.:** P5112
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5112 **SAS No.:** P5112
Matrix: Water **Level:** LOW **Client ID:** COMP-1A
Sample ID: P5136-02 **Spiked ID:** P5136-02A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Barium	ug/L	88 - 113	2160		1340		1000	82		P