

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5112 SAS No.: P5112 SDG NO.: P5112

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL01 Date Analyzed: 12/06/2024

Lab Sample No.: AR1660CCC500 Data File : PP068836.D Time Analyzed: 09:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.927	5.821	6.021	468.990	500.000	-6.2
Aroclor-1016-2	5.949	5.844	6.044	477.620	500.000	-4.5
Aroclor-1016-3	6.013	5.908	6.108	475.790	500.000	-4.8
Aroclor-1016-4	6.111	6.005	6.205	485.830	500.000	-2.8
Aroclor-1016-5	6.406	6.299	6.499	482.270	500.000	-3.5
Aroclor-1260-1	7.529	7.423	7.623	455.120	500.000	-9.0
Aroclor-1260-2	7.782	7.676	7.876	450.910	500.000	-9.8
Aroclor-1260-3	8.144	8.038	8.238	444.480	500.000	-11.1
Aroclor-1260-4	8.383	8.276	8.476	437.220	500.000	-12.6
Aroclor-1260-5	8.720	8.613	8.813	436.550	500.000	-12.7
Decachlorobiphenyl	10.681	10.572	10.772	46.900	50.000	-6.2
Tetrachloro-m-xylene	4.762	4.657	4.857	50.490	50.000	1.0

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5112 SAS No.: P5112 SDG NO.: P5112

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL01 Date Analyzed: 12/06/2024

Lab Sample No.: AR1660CCC500 Data File : PP068836.D Time Analyzed: 09:34

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.156	5.058	5.258	524.070	500.000	4.8
Aroclor-1016-2	5.175	5.079	5.279	511.580	500.000	2.3
Aroclor-1016-3	5.356	5.259	5.459	499.540	500.000	-0.1
Aroclor-1016-4	5.396	5.298	5.498	513.170	500.000	2.6
Aroclor-1016-5	5.614	5.517	5.717	506.480	500.000	1.3
Aroclor-1260-1	6.657	6.560	6.760	481.100	500.000	-3.8
Aroclor-1260-2	6.844	6.747	6.947	473.610	500.000	-5.3
Aroclor-1260-3	7.001	6.905	7.105	484.530	500.000	-3.1
Aroclor-1260-4	7.475	7.378	7.578	495.550	500.000	-0.9
Aroclor-1260-5	7.714	7.617	7.817	480.190	500.000	-4.0
Decachlorobiphenyl	9.204	9.110	9.310	48.300	50.000	-3.4
Tetrachloro-m-xylene	4.050	3.951	4.151	52.700	50.000	5.4