

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5112 SAS No.: P5112 SDG NO.: P5112

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 12/06/2024

Lab Sample No.: AR1660CCC500 Data File : PP068851.D Time Analyzed: 16:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.924	5.821	6.021	483.800	500.000	-3.2
Aroclor-1016-2	5.946	5.844	6.044	477.740	500.000	-4.5
Aroclor-1016-3	6.009	5.908	6.108	526.230	500.000	5.2
Aroclor-1016-4	6.108	6.005	6.205	533.260	500.000	6.7
Aroclor-1016-5	6.402	6.299	6.499	446.380	500.000	-10.7
Aroclor-1260-1	7.526	7.423	7.623	446.300	500.000	-10.7
Aroclor-1260-2	7.779	7.676	7.876	453.310	500.000	-9.3
Aroclor-1260-3	8.141	8.038	8.238	458.060	500.000	-8.4
Aroclor-1260-4	8.379	8.276	8.476	430.240	500.000	-14.0
Aroclor-1260-5	8.717	8.613	8.813	437.760	500.000	-12.4
Decachlorobiphenyl	10.675	10.572	10.772	46.710	50.000	-6.6
Tetrachloro-m-xylene	4.759	4.657	4.857	50.840	50.000	1.7

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5112 SAS No.: P5112 SDG NO.: P5112

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 12/06/2024

Lab Sample No.: AR1660CCC500 Data File : PP068851.D Time Analyzed: 16:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.152	5.058	5.258	510.620	500.000	2.1
Aroclor-1016-2	5.172	5.079	5.279	504.910	500.000	1.0
Aroclor-1016-3	5.352	5.259	5.459	501.150	500.000	0.2
Aroclor-1016-4	5.392	5.298	5.498	512.610	500.000	2.5
Aroclor-1016-5	5.610	5.517	5.717	491.840	500.000	-1.6
Aroclor-1260-1	6.653	6.560	6.760	469.080	500.000	-6.2
Aroclor-1260-2	6.840	6.747	6.947	459.980	500.000	-8.0
Aroclor-1260-3	6.998	6.905	7.105	466.380	500.000	-6.7
Aroclor-1260-4	7.471	7.378	7.578	471.870	500.000	-5.6
Aroclor-1260-5	7.710	7.617	7.817	460.920	500.000	-7.8
Decachlorobiphenyl	9.200	9.110	9.310	45.310	50.000	-9.4
Tetrachloro-m-xylene	4.047	3.951	4.151	52.080	50.000	4.2