

Surrogate Summary

SDG No.: **P5112**

Client: **Tully Construction Co., Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP068360.D	PIBLK-PP068360.D	Tetrachloro-m-xylene	1	20	21.0	105		60	140
		Decachlorobiphenyl	1	20	23.4	117		60	140
		Tetrachloro-m-xylene	2	20	21.6	108		60	140
		Decachlorobiphenyl	2	20	23.6	118		60	140
I.BLK-PP068840.D	PIBLK-PP068840.D	Tetrachloro-m-xylene	1	20	20.0	100		60	140
		Decachlorobiphenyl	1	20	21.5	107		60	140
		Tetrachloro-m-xylene	2	20	20.8	104		60	140
		Decachlorobiphenyl	2	20	21.0	105		60	140
PB165421BL	PB165421BL	Tetrachloro-m-xylene	1	20	18.6	93		32	144
		Decachlorobiphenyl	1	20	19.6	98		32	175
		Tetrachloro-m-xylene	2	20	19.5	98		32	144
		Decachlorobiphenyl	2	20	19.7	98		32	175
PB165421BS	PB165421BS	Tetrachloro-m-xylene	1	20	19.6	98		32	144
		Decachlorobiphenyl	1	20	15.6	78		32	175
		Tetrachloro-m-xylene	2	20	20.3	102		32	144
		Decachlorobiphenyl	2	20	20.5	103		32	175
P5112-01	10TH-ST-SOIL	Tetrachloro-m-xylene	1	20	19.3	96		32	144
		Decachlorobiphenyl	1	20	19.8	99		32	175
		Tetrachloro-m-xylene	2	20	21.2	106		32	144
		Decachlorobiphenyl	2	20	19.6	98		32	175
P5117-02MS	TAPIAL2-IDW-SOIL-120424-00-T2	Tetrachloro-m-xylene	1	20	20.7	103		32	144
		Decachlorobiphenyl	1	20	16.8	84		32	175
		Tetrachloro-m-xylene	2	20	19.9	100		32	144
		Decachlorobiphenyl	2	20	17.6	88		32	175
P5117-02MSD	TAPIAL2-IDW-SOIL-120424-00-T2	Tetrachloro-m-xylene	1	20	20.1	100		32	144
		Decachlorobiphenyl	1	20	17.1	85		32	175
		Tetrachloro-m-xylene	2	20	19.6	98		32	144
		Decachlorobiphenyl	2	20	17.3	87		32	175
I.BLK-PP068855.D	PIBLK-PP068855.D	Tetrachloro-m-xylene	1	20	21.4	107		60	140
		Decachlorobiphenyl	1	20	21.3	106		60	140
		Tetrachloro-m-xylene	2	20	20.5	103		60	140
		Decachlorobiphenyl	2	20	21.3	107		60	140