

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5112

Client: Tully Construction Co., Inc.

Analytical Method: 8082A

DataFile : PP068846.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID:	TAPIAL2-IDW-SOIL-120424-00-T2MSD											
P5117-02MSD	AR1016	208.6	0	221	ug/kg	106		2		55	146	20
	AR1260	208.6	0	205	ug/kg	98		1		45	144	20