

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: TULL02
 ProjectID: 10th Street & 2nd Avenue
 Lab Code: CHEM Case No.: P5112 SAS No.: P5112 SDG No.: P5112

Calibration Sequence : FG112024		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	191243670	112496	FG014888.D
850	93606491	110125	FG014889.D
340	38457325	113110	FG014890.D
170	21518915	126582	FG014891.D
85	10451945	122964	FG014892.D
AVG RF : 117055		% RSD : 6.191	AVG RT : 15.0408