

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: TULL02
ProjectID: 10th Street & 2nd Avenue
Lab Code: CHEM Case No.: P5112 SAS No.: P5112 SDG No.: P5112
DataFile: FG014925.D Analyst Name: YP\AJ Analyst Date: 12-06-2024

Conc. (PPM)	Area Count	RF	Average RF	%D
850	92827022	109208	117055	6.704