

SOIL TPH GC SURROGATE RECOVERY

Lab Name: Chemtech Client: Tully Construction Co., Inc.
Lab Code: CHEM Case No.: P5112 SAS No.: P5112 SDG No.: P5112

EPA SAMPLE NO.	S1 TETRACOSANE-d50	S2	S3	S4	TOT OUT
PIBLK-FG014924.D	79				0
PIBLK-FG014932.D	92				0
PIBLK-FG014935.D	82				0
PIBLK-FG014939.D	85				0
10TH-ST-SOIL	91				0
10TH-ST-SOILMS	68				0
10TH-ST-SOILMSD	69				0
PB165424BL	80				0
PB165424BS	88				0

QC LIMITS

TETRACOSANE-d50

For Water : 29-130
For Soil : 37-130

Column to be used to flag recovery values
* Values outside of contract required QC limits
D Surrogate Diluted Out