

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TULL02
 Lab Code: CHEM Case No.: P5112 SAS No.: P5112 SDG No.: P5112
 Instrument ID: MSVOA_Y Calibration Date(s): 12/02/2024 12/02/2024
 Heated Purge: (Y/N) Y Calibration Time(s): 09:16 11:09
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF005 = VY020472.D			RRF010 = VY020473.D		RRF020 = VY020474.D			
	RRF050 = VY020475.D			RRF100 = VY020476.D		RRF150 = VY020477.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Dichlorodifluoromethane	0.506	0.450	0.422	0.527	0.509	0.512	0.488	8.5	
Chloromethane	0.534	0.456	0.433	0.525	0.497	0.518	0.494	8.2	
Vinyl Chloride	0.526	0.476	0.443	0.540	0.511	0.526	0.503	7.4	
Bromomethane	0.322	0.285	0.254	0.307	0.298	0.306	0.295	8	
Chloroethane	0.325	0.298	0.285	0.317	0.308	0.318	0.309	4.8	
Trichlorofluoromethane	1.019	0.919	0.858	0.971	0.929	0.962	0.943	5.8	
1,1,2-Trichlorotrifluoroethane	0.615	0.548	0.530	0.551	0.516	0.537	0.549	6.3	
1,1-Dichloroethene	0.560	0.505	0.472	0.534	0.511	0.532	0.519	5.8	
Acetone	0.134	0.111	0.099	0.146	0.126	0.142	0.126	14.4	
Carbon Disulfide	1.446	1.331	1.244	1.660	1.588	1.645	1.486	11.7	
Methyl tert-butyl Ether	1.468	1.275	1.213	1.358	1.273	1.366	1.326	6.8	
Methyl Acetate	0.302	0.270	0.235	0.279	0.248	0.279	0.269	9	
Methylene Chloride	0.588	0.536	0.499	0.537	0.524	0.544	0.538	5.4	
trans-1,2-Dichloroethene	0.648	0.565	0.528	0.580	0.564	0.573	0.576	6.8	
1,1-Dichloroethane	1.215	1.144	1.053	1.092	1.058	1.089	1.108	5.5	
Cyclohexane	1.172	1.028	0.912	0.999	0.949	0.951	1.002	9.3	
2-Butanone	0.171	0.149	0.137	0.179	0.151	0.171	0.160	10.2	
Carbon Tetrachloride	0.670	0.639	0.600	0.649	0.619	0.629	0.634	3.8	
cis-1,2-Dichloroethene	0.723	0.677	0.644	0.675	0.648	0.672	0.673	4.2	
Bromochloromethane	0.447	0.385	0.364	0.445	0.419	0.423	0.414	8.1	
Chloroform	1.268	1.164	1.086	1.111	1.085	1.111	1.137	6.2	
1,1,1-Trichloroethane	1.217	1.120	1.030	1.084	1.055	1.066	1.095	6.1	
Methylcyclohexane	0.669	0.643	0.602	0.693	0.676	0.674	0.660	4.9	
Benzene	1.687	1.588	1.474	1.586	1.531	1.538	1.567	4.6	
1,2-Dichloroethane	0.436	0.414	0.381	0.428	0.397	0.415	0.412	4.9	
Trichloroethene	0.409	0.387	0.373	0.388	0.379	0.381	0.386	3.2	
1,2-Dichloropropane	0.414	0.371	0.354	0.360	0.348	0.353	0.367	6.7	
Bromodichloromethane	0.588	0.539	0.507	0.531	0.515	0.531	0.535	5.3	
4-Methyl-2-Pentanone	0.222	0.206	0.190	0.225	0.199	0.218	0.210	6.7	
Toluene	1.027	0.984	0.924	0.995	0.963	0.974	0.978	3.5	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TULL02
Lab Code: CHEM Case No.: P5112 SAS No.: P5112 SDG No.: P5112
Instrument ID: MSVOA_Y Calibration Date(s): 12/02/2024 12/02/2024
Heated Purge: (Y/N) Y Calibration Time(s): 09:16 11:09
GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY020472.D		RRF010 = VY020473.D		RRF020 = VY020474.D			
		RRF050 = VY020475.D		RRF100 = VY020476.D		RRF150 = VY020477.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
t-1,3-Dichloropropene	0.503	0.463	0.445	0.496	0.477	0.497	0.480	4.7	
cis-1,3-Dichloropropene	0.595	0.556	0.534	0.581	0.563	0.577	0.568	3.8	
1,1,2-Trichloroethane	0.258	0.252	0.227	0.247	0.231	0.238	0.242	4.9	
2-Hexanone	0.154	0.140	0.131	0.172	0.149	0.163	0.152	9.9	
Dibromochloromethane	0.371	0.336	0.311	0.346	0.323	0.337	0.337	6	
1,2-Dibromoethane	0.245	0.215	0.207	0.227	0.213	0.223	0.222	6.1	
Tetrachloroethene	0.456	0.434	0.400	0.420	0.408	0.417	0.422	4.8	
Chlorobenzene	1.336	1.297	1.200	1.233	1.213	1.240	1.253	4.2	
Ethyl Benzene	2.481	2.386	2.268	2.344	2.291	2.325	2.349	3.2	
m/p-Xylenes	0.909	0.881	0.837	0.848	0.839	0.847	0.860	3.3	
o-Xylene	0.841	0.834	0.779	0.805	0.780	0.796	0.806	3.3	
Styrene	1.403	1.379	1.293	1.347	1.311	1.334	1.344	3.1	
Bromoform	0.248	0.224	0.211	0.230	0.213	0.230	0.226	5.9	
Isopropylbenzene	5.309	4.936	4.771	4.571	4.664	4.775	4.838	5.4	
1,1,2,2-Tetrachloroethane	0.794	0.697	0.662	0.685	0.644	0.692	0.696	7.5	
1,3-Dichlorobenzene	2.129	1.990	1.872	1.846	1.850	1.905	1.932	5.7	
1,4-Dichlorobenzene	2.082	1.949	1.853	1.809	1.797	1.859	1.892	5.7	
1,2-Dichlorobenzene	1.772	1.698	1.595	1.579	1.548	1.639	1.639	5.1	
1,2-Dibromo-3-Chloropropane	0.112	0.106	0.096	0.108	0.099	0.111	0.105	6.4	
1,2,4-Trichlorobenzene	0.919	0.890	0.825	0.921	0.965	1.007	0.921	6.8	
1,2,3-Trichlorobenzene	0.715	0.722	0.666	0.759	0.782	0.818	0.743	7.3	
1,2-Dichloroethane-d4	0.582	0.471	0.439	0.557	0.491	0.511	0.509	10.5	
Dibromofluoromethane	0.352	0.296	0.275	0.346	0.308	0.311	0.315	9.4	
Toluene-d8	1.315	1.127	1.059	1.320	1.183	1.188	1.199	8.6	
4-Bromofluorobenzene	0.504	0.389	0.364	0.451	0.403	0.405	0.419	12	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.