

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5112

Client: Tully Construction Co., Inc.

Analytical Method: SW8260D

Datafile : VY020531.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1206SBSD01	Dichlorodifluoromethane	20	18.6	ug/Kg	93	1		64	136	20
	Chloromethane	20	17.8	ug/Kg	89	3		70	130	20
	Vinyl chloride	20	17.7	ug/Kg	89	0		72	129	20
	Bromomethane	20	18.6	ug/Kg	93	4		58	141	20
	Chloroethane	20	18.5	ug/Kg	93	2		69	130	20
	Trichlorofluoromethane	20	19.3	ug/Kg	97	0		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	20.4	ug/Kg	102	0		81	123	20
	1,1-Dichloroethene	20	19.6	ug/Kg	98	3		79	121	20
	Acetone	100	100	ug/Kg	100	3		60	131	20
	Carbon disulfide	20	16.4	ug/Kg	82	0		45	154	20
	Methyl tert-butyl Ether	20	23.3	ug/Kg	117	5		77	129	20
	Methyl Acetate	20	25.9	ug/Kg	130	7		69	149	20
	Methylene Chloride	20	22.5	ug/Kg	113	6		56	174	20
	trans-1,2-Dichloroethene	20	19.8	ug/Kg	99	3		80	123	20
	1,1-Dichloroethane	20	20.8	ug/Kg	104	0		82	123	20
	Cyclohexane	20	18.6	ug/Kg	93	1		76	122	20
	2-Butanone	100	120	ug/Kg	120	9		69	131	20
	Carbon Tetrachloride	20	20.4	ug/Kg	102	2		76	129	20
	cis-1,2-Dichloroethene	20	21.2	ug/Kg	106	2		82	123	20
	Bromochloromethane	20	20.5	ug/Kg	103	4		80	127	20
	Chloroform	20	21.8	ug/Kg	109	3		82	125	20
	1,1,1-Trichloroethane	20	20.7	ug/Kg	104	1		80	126	20
	Methylcyclohexane	20	18.3	ug/Kg	92	0		77	123	20
	Benzene	20	20.4	ug/Kg	102	1		84	121	20
	1,2-Dichloroethane	20	22.5	ug/Kg	113	4		81	126	20
	Trichloroethene	20	20.4	ug/Kg	102	1		83	122	20
	1,2-Dichloropropane	20	20.9	ug/Kg	104	2		83	122	20
	Bromodichloromethane	20	21.9	ug/Kg	110	4		82	123	20
	4-Methyl-2-Pentanone	100	130	ug/Kg	130	8		70	135	20
	Toluene	20	20.6	ug/Kg	103	1		83	122	20
	t-1,3-Dichloropropene	20	22.2	ug/Kg	111	2		78	124	20
	cis-1,3-Dichloropropene	20	21.4	ug/Kg	107	3		81	122	20
	1,1,2-Trichloroethane	20	23.5	ug/Kg	117	3		82	125	20
	2-Hexanone	100	120	ug/Kg	120	9		66	138	20
	Dibromochloromethane	20	23.3	ug/Kg	117	5		79	125	20
	1,2-Dibromoethane	20	23.1	ug/Kg	116	6		80	125	20
	Tetrachloroethene	20	20.5	ug/Kg	103	4		83	125	20
	Chlorobenzene	20	20.7	ug/Kg	104	2		84	122	20
	Ethyl Benzene	20	19.7	ug/Kg	99	0		82	124	20
	m/p-Xylenes	40	39.9	ug/Kg	100	1		83	124	20
	o-Xylene	20	20.4	ug/Kg	102	0		83	123	20
	Styrene	20	20.8	ug/Kg	104	0		82	124	20
	Bromoform	20	23.3	ug/Kg	117	4		75	127	20
	Isopropylbenzene	20	18.7	ug/Kg	94	3		82	124	20
	1,1,2,2-Tetrachloroethane	20	21.0	ug/Kg	105	2		77	127	20
	1,3-Dichlorobenzene	20	20.5	ug/Kg	103	5		83	122	20
	1,4-Dichlorobenzene	20	20.3	ug/Kg	102	0		84	121	20
	1,2-Dichlorobenzene	20	20.9	ug/Kg	104	0		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5112
Client: Tully Construction Co., Inc.
Analytical Method: SW8260D **Datafile :** VY020531.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1206SBSD01	1,2-Dibromo-3-Chloropropane	20	23.2	ug/Kg	116	6		66	134	20
	1,2,4-Trichlorobenzene	20	22.0	ug/Kg	110	6		78	127	20
	1,2,3-Trichlorobenzene	20	23.2	ug/Kg	116	5		70	137	20