

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL02
 Lab Code: CHEM Case No.: P5112 SAS No.: P5112 SDG NO.: P5112
 Lab File ID: VY020528.D Date Analyzed: 12/06/2024
 Instrument ID: MSVOA_Y Time Analyzed: 10:54
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	98467	13.352				
UPPER LIMIT	196934	13.852				
LOWER LIMIT	49233.5	12.852				
EPA SAMPLE NO.						
10TH-ST-SOIL	54408	13.35				
VY1206SBL01	85714	13.35				
VY1206SBS01	94820	13.35				
VY1206SBSD01	95244	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.