

Initial and Continuing Calibration Verification

Client: Tetra Tech NUS, Inc.

SDG No.: P5113

Project: NWIRP Calverton Site 16 Phase 3 RI

RunNo.: LB133779

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: TOC	ICV1	mg/L	934	1000	93	90-110	11/12/2024
Sample ID: TOC	CCV1	mg/L	937	1000	94	90-110	12/06/2024
Sample ID: TOC	CCV2	mg/L	999	1000	100	90-110	12/06/2024