



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: Tetra Tech NUS, Inc.

SDG No.: P5113

Project: NWIRP Calverton Site 16 Phase 3 RI

RunNo.: LB133779

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TOC	mg/L	< 125.0000	125.0000	U	22.3	250	11/12/2024
Sample ID: CCB1 TOC	mg/L	< 125.0000	125.0000	U	22.3	250	12/06/2024
Sample ID: CCB2 TOC	mg/L	< 125.0000	125.0000	U	22.3	250	12/06/2024