ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEM	TECH PROJECT NUMBER: P5117	MATRIX: TCLP		
METH	OD: 8270E/3510/1311			
1.	Chromatograms Labeled/Compounds Identified. (Field samples	NA s and Method Blanks)	NO	YES ✔
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR N CLP AND NJ)	Y ASP CLP,		~
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 s 8000 Series.	series and 12 hours for		✓
4.	GC/MS Calibration - Initial Calibration performed within 30 day analysis and continuing calibration performed within 24 hours of 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.			√
	The Initial Calibration met the requirements . The Continuous Calibration met the requirements .			
6.	Blank Contamination - If yes, list compounds and concentration	is in each blank:	\checkmark	
7.	Surrogate Recoveries Meet Criteria			√
	If not met, list those compounds and their recoveries which fall ranges.	outside the acceptable		
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria			\checkmark
	If not met, list those compounds and their recoveries which fall range.	outside the acceptable		
	The Blank Spike met requirements for all samples .			
	The RPD for {P5095-04MSD} with File ID: BF140792.D met of Nitrobenzene[22%] due to difference in results of MS and MSD			
9.	Internal Standard Area/Retention Time Shift Meet Criteria			\checkmark
	Comments:			

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

<u>GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY</u> (CONTINUED)

		NA	NO	YES
10.	Extraction Holding Time Met			\checkmark
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The Form 6 is not included in the data package because the Initial Calibration was performed using 8 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date