

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133776

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/06/2024	08:55
2	CAL2	1	Water	NA	NA	20.2	7.00	12/06/2024	08:56
3	CAL3	1	Water	NA	NA	20.3	10.02	12/06/2024	08:58
4	ICV	1	Water	NA	NA	20.3	7.00	12/06/2024	09:00
5	CCV1	1	Water	NA	NA	20.4	2.02	12/06/2024	09:20
6	P5117-01	1	Solid	20.02	20	20.8	7.21	12/06/2024	09:30
7	P5117-01DUP	1	Solid	20.03	20	20.9	7.22	12/06/2024	09:31
8	P5117-02	1	Solid	20.04	20	20.1	6.41	12/06/2024	09:35
9	CCV2	1	Water	NA	NA	20.3	12.02	12/06/2024	09:37

WORKLIST(Hardcopy Internal Chain)

133776

WorkList Name : PH P5117

WorkList ID : 186055

Department : Wet-Chemistry

Date : 12-06-2024 08:44:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5117-01	TAPIAL3-SB04I-10-120324-00-	Solid	pH	Cool 4 deg C	WEST04	L41	12/05/2024	9045D
P5117-02	TAPIAL2-IDW-SOIL-120424-00-	Solid	pH	Cool 4 deg C	WEST04	L41	12/05/2024	9045D

Date/Time 12/06/24 08:50

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 12/06/24

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]