



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Initial and Continuing Calibration Verification

**Client:** Weston Solutions

**SDG No.:** P5117

**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**RunNo.:** LB133776

| Analyte    |      | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|------------|------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID: | ICV  |       |        |            |               |                           |                  |
| pH         |      | pH    | 7.00   | 7          | 100           | 90-110                    | 12/06/2024       |
| Sample ID: | CCV1 |       |        |            |               |                           |                  |
| pH         |      | pH    | 2.02   | 2.00       | 101           | 90-110                    | 12/06/2024       |
| Sample ID: | CCV2 |       |        |            |               |                           |                  |
| pH         |      | pH    | 12.02  | 12.00      | 100           | 90-110                    | 12/06/2024       |

## Initial and Continuing Calibration Verification

**Client:** Weston Solutions

**SDG No.:** P5117

**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**RunNo.:** LB133779

| Analyte    |      | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|------------|------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID: | ICV1 |       |        |            |               |                           |                  |
| TOC        |      | mg/L  | 934    | 1000       | 93            | 90-110                    | 11/12/2024       |
| Sample ID: | CCV1 |       |        |            |               |                           |                  |
| TOC        |      | mg/L  | 937    | 1000       | 94            | 90-110                    | 12/06/2024       |
| Sample ID: | CCV2 |       |        |            |               |                           |                  |
| TOC        |      | mg/L  | 999    | 1000       | 100           | 90-110                    | 12/06/2024       |

## Initial and Continuing Calibration Verification

**Client:** Weston Solutions

**SDG No.:** P5117

**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**RunNo.:** LB133847

| Analyte    |      | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|------------|------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: | ICV1 |       |        |            |            |                        |               |
| Cyanide    |      | mg/L  | 0.1    | 0.099      | 101        | 90-110                 | 12/09/2024    |
| Sample ID: | CCV1 |       |        |            |            |                        |               |
| Cyanide    |      | mg/L  | 0.26   | 0.25       | 104        | 90-110                 | 12/09/2024    |
| Sample ID: | CCV2 |       |        |            |            |                        |               |
| Cyanide    |      | mg/L  | 0.25   | 0.25       | 100        | 90-110                 | 12/09/2024    |
| Sample ID: | CCV3 |       |        |            |            |                        |               |
| Cyanide    |      | mg/L  | 0.27   | 0.25       | 108        | 90-110                 | 12/09/2024    |