



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Weston Solutions</u>				SDG No.:	<u>P5117</u>				
Contract:	<u>WEST04</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P5117</u>	SAS No.:	<u>P5117</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB12	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/06/2024	16:44	LB133794

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions

SDG No.: P5117

Contract: WEST04

Lab Code: CHEM

Case No.: P5117

SAS No.: P5117

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/06/2024	16:51	LB133794
CCB02	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/06/2024	17:37	LB133794
CCB03	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/06/2024	18:07	LB133794
CCB04	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/06/2024	18:29	LB133794

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions SDG No.: P5117
Contract: WEST04 Lab Code: CHEM Case No.: P5117 SAS No.: P5117

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	20.0	+/-20.0	U	10.0	20.0	P	12/19/2024	16:07	LB134024
	Antimony	2.00	+/-2.00	U	0.25	2.00	P	12/19/2024	16:07	LB134024
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	16:07	LB134024
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/19/2024	16:07	LB134024
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	16:07	LB134024
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/19/2024	16:07	LB134024
	Calcium	500	+/-500	U	190	500	P	12/19/2024	16:07	LB134024
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/19/2024	16:07	LB134024
	Cobalt	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	16:07	LB134024
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/19/2024	16:07	LB134024
	Iron	50.0	+/-50.0	U	25.0	50.0	P	12/19/2024	16:07	LB134024
	Lead	1.00	+/-1.00	U	0.75	1.00	P	12/19/2024	16:07	LB134024
	Magnesium	500	+/-500	U	190	500	P	12/19/2024	16:07	LB134024
	Manganese	1.00	+/-1.00	U	0.75	1.00	P	12/19/2024	16:07	LB134024
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	16:07	LB134024
	Potassium	500	+/-500	U	190	500	P	12/19/2024	16:07	LB134024
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/19/2024	16:07	LB134024
	Silver	1.00	+/-1.00	U	0.50	1.00	P	12/19/2024	16:07	LB134024
	Sodium	500	+/-500	U	190	500	P	12/19/2024	16:07	LB134024
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	12/19/2024	16:07	LB134024
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/19/2024	16:07	LB134024
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/19/2024	16:07	LB134024

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>P5117</u>								
Contract: <u>WEST04</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5117</u>			SAS No.: <u>P5117</u>					
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	20.0	+/-20.0	U	10.0	20.0	P	12/19/2024	16:19	LB134024
	Antimony	0.19	+/-2.00	J	0.25	2.00	P	12/19/2024	16:19	LB134024
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	16:19	LB134024
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/19/2024	16:19	LB134024
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	16:19	LB134024
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/19/2024	16:19	LB134024
	Calcium	500	+/-500	U	190	500	P	12/19/2024	16:19	LB134024
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/19/2024	16:19	LB134024
	Cobalt	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	16:19	LB134024
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/19/2024	16:19	LB134024
	Iron	50.0	+/-50.0	U	25.0	50.0	P	12/19/2024	16:19	LB134024
	Lead	0.31	+/-1.00	J	0.75	1.00	P	12/19/2024	16:19	LB134024
	Magnesium	500	+/-500	U	190	500	P	12/19/2024	16:19	LB134024
	Manganese	1.00	+/-1.00	U	0.75	1.00	P	12/19/2024	16:19	LB134024
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	16:19	LB134024
	Potassium	500	+/-500	U	190	500	P	12/19/2024	16:19	LB134024
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/19/2024	16:19	LB134024
	Silver	0.080	+/-1.00	J	0.50	1.00	P	12/19/2024	16:19	LB134024
	Sodium	500	+/-500	U	190	500	P	12/19/2024	16:19	LB134024
	Thallium	0.14	+/-1.00	J	0.50	1.00	P	12/19/2024	16:19	LB134024
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/19/2024	16:19	LB134024
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/19/2024	16:19	LB134024
CCB02	Aluminum	3.07	+/-20.0	J	10.0	20.0	P	12/19/2024	17:03	LB134024
	Antimony	0.25	+/-2.00	J	0.25	2.00	P	12/19/2024	17:03	LB134024
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	17:03	LB134024
	Barium	0.35	+/-10.0	J	1.25	10.0	P	12/19/2024	17:03	LB134024
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	17:03	LB134024
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/19/2024	17:03	LB134024
	Calcium	500	+/-500	U	190	500	P	12/19/2024	17:03	LB134024
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/19/2024	17:03	LB134024
	Cobalt	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	17:03	LB134024
	Copper	0.43	+/-2.00	J	1.50	2.00	P	12/19/2024	17:03	LB134024
	Iron	50.0	+/-50.0	U	25.0	50.0	P	12/19/2024	17:03	LB134024
	Lead	0.39	+/-1.00	J	0.75	1.00	P	12/19/2024	17:03	LB134024
	Magnesium	500	+/-500	U	190	500	P	12/19/2024	17:03	LB134024
	Manganese	0.35	+/-1.00	J	0.75	1.00	P	12/19/2024	17:03	LB134024
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	17:03	LB134024
	Potassium	500	+/-500	U	190	500	P	12/19/2024	17:03	LB134024
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/19/2024	17:03	LB134024

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>P5117</u>								
Contract: <u>WEST04</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5117</u>			SAS No.: <u>P5117</u>					
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	0.11	+/-1.00	J	0.50	1.00	P	12/19/2024	17:03	LB134024
	Sodium	500	+/-500	U	190	500	P	12/19/2024	17:03	LB134024
	Thallium	0.11	+/-1.00	J	0.50	1.00	P	12/19/2024	17:03	LB134024
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/19/2024	17:03	LB134024
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/19/2024	17:03	LB134024
CCB03	Aluminum	2.64	+/-20.0	J	10.0	20.0	P	12/19/2024	17:23	LB134024
	Antimony	0.23	+/-2.00	J	0.25	2.00	P	12/19/2024	17:23	LB134024
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	17:23	LB134024
	Barium	0.35	+/-10.0	J	1.25	10.0	P	12/19/2024	17:23	LB134024
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	17:23	LB134024
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/19/2024	17:23	LB134024
	Calcium	500	+/-500	U	190	500	P	12/19/2024	17:23	LB134024
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/19/2024	17:23	LB134024
	Cobalt	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	17:23	LB134024
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/19/2024	17:23	LB134024
	Iron	50.0	+/-50.0	U	25.0	50.0	P	12/19/2024	17:23	LB134024
	Lead	0.38	+/-1.00	J	0.75	1.00	P	12/19/2024	17:23	LB134024
	Magnesium	500	+/-500	U	190	500	P	12/19/2024	17:23	LB134024
	Manganese	0.26	+/-1.00	J	0.75	1.00	P	12/19/2024	17:23	LB134024
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/19/2024	17:23	LB134024
	Potassium	500	+/-500	U	190	500	P	12/19/2024	17:23	LB134024
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/19/2024	17:23	LB134024
	Silver	0.11	+/-1.00	J	0.50	1.00	P	12/19/2024	17:23	LB134024
	Sodium	500	+/-500	U	190	500	P	12/19/2024	17:23	LB134024
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	12/19/2024	17:23	LB134024
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/19/2024	17:23	LB134024
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/19/2024	17:23	LB134024