



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Weston Solutions</u>				SDG No.:	<u>P5117</u>				
Contract:	<u>WEST04</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P5117</u>	SAS No.:	<u>P5117</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB13	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/09/2024	11:48	LB133829

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions

SDG No.: P5117

Contract: WEST04

Lab Code: CHEM

Case No.: P5117

SAS No.: P5117

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/09/2024	11:53	LB133829
CCB06	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/09/2024	12:28	LB133829
CCB07	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/09/2024	12:55	LB133829
CCB08	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/09/2024	13:48	LB133829
CCB09	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/09/2024	15:35	LB133829
CCB10	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/09/2024	15:47	LB133829

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions **SDG No.:** P5117
Contract: WEST04 **Lab Code:** CHEM **Case No.:** P5117 **SAS No.:** P5117

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/09/2024	11:54	LB133834
	Barium	100	+/-100	U	25.0	100	P	12/09/2024	11:54	LB133834
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/09/2024	11:54	LB133834
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/09/2024	11:54	LB133834
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/09/2024	11:54	LB133834
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/09/2024	11:54	LB133834
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/09/2024	11:54	LB133834

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>P5117</u>								
Contract: <u>WEST04</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5117</u>	SAS No.: <u>P5117</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/09/2024	12:34	LB133834
	Barium	100	+/-100	U	25.0	100	P	12/09/2024	12:34	LB133834
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/09/2024	12:34	LB133834
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/09/2024	12:34	LB133834
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/09/2024	12:34	LB133834
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/09/2024	12:34	LB133834
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/09/2024	12:34	LB133834
CCB02	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/09/2024	13:35	LB133834
	Barium	100	+/-100	U	25.0	100	P	12/09/2024	13:35	LB133834
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/09/2024	13:35	LB133834
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/09/2024	13:35	LB133834
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/09/2024	13:35	LB133834
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/09/2024	13:35	LB133834
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/09/2024	13:35	LB133834
CCB03	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/09/2024	14:27	LB133834
	Barium	100	+/-100	U	25.0	100	P	12/09/2024	14:27	LB133834
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/09/2024	14:27	LB133834
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/09/2024	14:27	LB133834
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/09/2024	14:27	LB133834
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/09/2024	14:27	LB133834
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/09/2024	14:27	LB133834
CCB04	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/09/2024	15:17	LB133834
	Barium	100	+/-100	U	25.0	100	P	12/09/2024	15:17	LB133834
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/09/2024	15:17	LB133834
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/09/2024	15:17	LB133834
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/09/2024	15:17	LB133834
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/09/2024	15:17	LB133834
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/09/2024	15:17	LB133834
CCB05	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/09/2024	15:56	LB133834
	Barium	100	+/-100	U	25.0	100	P	12/09/2024	15:56	LB133834
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/09/2024	15:56	LB133834
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/09/2024	15:56	LB133834
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/09/2024	15:56	LB133834
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/09/2024	15:56	LB133834
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/09/2024	15:56	LB133834