

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions **SDG No.:** P5117
Contract: WEST04 **Lab Code:** CHEM **Case No.:** P5117 **SAS No.:** P5117

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165436BS							
Arsenic	ug/L	4000	3680		92	87 - 113	P
Barium	ug/L	1000	959		96	88 - 113	P
Cadmium	ug/L	1000	935		94	88 - 113	P
Chromium	ug/L	2000	1940		97	90 - 113	P
Lead	ug/L	5000	4660		93	86 - 113	P
Selenium	ug/L	10000	9370		94	83 - 114	P
Silver	ug/L	380	354		93	84 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions

SDG No.: P5117

Contract: WEST04

Lab Code: CHEM

Case No.: P5117

SAS No.: P5117

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165486BS Mercury	ug/L	4.0	3.40		85	82 - 119	CV