



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** P5117

**Client:** Weston Solutions

**Analytical Method:** **8082A** **Datafile :** PP068842.D

| Lab Sample ID | Parameter | Spike | Result | Units | Rec | RPD | Qual | RPD  |     | Limits |     |
|---------------|-----------|-------|--------|-------|-----|-----|------|------|-----|--------|-----|
|               |           |       |        |       |     |     |      | Qual | Low | High   | RPD |
| PB165421BS    | AR1016    | 166.6 | 150    | ug/kg | 90  |     |      |      | 47  | 134    |     |
|               | AR1260    | 166.6 | 135    | ug/kg | 81  |     |      |      | 53  | 140    |     |