



Contract:	<u>WEST04</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>P5117</u>	SAS No.:	<u>P5117</u>	SDG NO.:	<u>P5117</u>
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):		<u>11/26/2024</u>	<u>11/26/2024</u>		
		Calibration Times:		<u>12:48</u>	<u>14:25</u>		

LAB FILE ID:		RT 200 =	<u>PS028632.D</u>	RT 500 =	<u>PS028633.D</u>
RT 750 =	<u>PS028634.D</u>	RT 1000 =	<u>PS028635.D</u>	RT 1500 =	<u>PS028636.D</u>

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: WEST04

Lab Code: CHEM Case No.: P5117 SAS No.: P5117 SDG NO.: P5117

Instrument ID: ECD_S Calibration Date(s): 11/26/2024 11/26/2024

Calibration Times: 12:48 14:25

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS028632.D</u>	RT 500 = <u>PS028633.D</u>
	RT 750 = <u>PS028634.D</u>	RT 1000 = <u>PS028635.D</u>
		RT 1500 = <u>PS028636.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW FROM TO	
2,4,5-TP(Silvex)	9.87	9.87	9.87	9.87	9.87	9.87	9.77	9.97
2,4-D	8.97	8.97	8.97	8.97	8.97	8.97	8.87	9.07
2,4-DCAA	7.73	7.72	7.72	7.73	7.73	7.72	7.62	7.82