

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: WEST04

Lab Code: CHEM Case No.: P5117 SAS No.: P5117 SDG NO.: P5117

Instrument ID: ECD_S Calibration Date(s): 11/26/2024 11/26/2024
Calibration Times: 12:48 14:25

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028632.D</u>	CF 500 =	<u>PS028633.D</u>		
CF 750 =		<u>PS028634.D</u>	CF 1000 =	<u>PS028635.D</u>	CF 1500 =	<u>PS028636.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	20271400000	18910000000	18123500000	17499500000	16594600000	18279800000	8
2,4-D	3843050000	3484790000	3301520000	3180130000	3051260000	3372150000	9
2,4-DCAA	2980930000	2776190000	2647410000	2550090000	2457090000	2682340000	8

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: WEST04
Lab Code: CHEM **Case No.:** P5117 **SAS No.:** P5117 **SDG NO.:** P5117
Instrument ID: ECD_S **Calibration Date(s):** 11/26/2024 11/26/2024
Calibration Times: 12:48 14:25

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 = <u>PS028632.D</u>		CF 500 = <u>PS028633.D</u>			
CF 750 = <u>PS028634.D</u>		CF 1000 = <u>PS028635.D</u>		CF 1500 = <u>PS028636.D</u>			
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	9649960000	9698620000	9598260000	9499290000	9266470000	9542520000	2
2,4-D	1782570000	1704130000	1676500000	1652880000	1638760000	1690970000	3
2,4-DCAA	1414130000	1352070000	1334210000	1323400000	1311340000	1347030000	3