

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM **Case No.:** P5117 **SAS No.:** P5117 **SDG NO.:** P5117

Continuing Calib Date: 12/06/2024 **Initial Calibration Date(s):** 11/26/2024 11/26/2024

Continuing Calib Time: 12:37 **Initial Calibration Time(s):** 12:48 14:25

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.23	7.23	7.13	7.33	0.00
2,4-D	8.35	8.35	8.25	8.45	0.00
2,4,5-TP(Silvex)	9.23	9.23	9.13	9.33	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM **Case No.:** P5117 **SAS No.:** P5117 **SDG NO.:** P5117

Continuing Calib Date: 12/06/2024 **Initial Calibration Date(s):** 11/26/2024 11/26/2024

Continuing Calib Time: 12:37 **Initial Calibration Time(s):** 12:48 14:25

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.72	7.72	7.62	7.82	0.00
2,4-D	8.96	8.97	8.87	9.07	0.01
2,4,5-TP(Silvex)	9.87	9.87	9.77	9.97	0.01