

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5117 SAS No.: P5117 SDG NO.: P5117

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/26/2024 11/26/2024

Client Sample No.: CCAL02 Date Analyzed: 12/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028727.D Time Analyzed: 19:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.224	9.128	9.328	760.570	712.500	6.7
2,4-D	8.348	8.251	8.451	732.590	705.000	3.9
2,4-DCAA	7.225	7.126	7.326	752.340	750.000	0.3

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5117 SAS No.: P5117 SDG NO.: P5117

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/26/2024 11/26/2024

Client Sample No.: CCAL02 Date Analyzed: 12/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028727.D Time Analyzed: 19:20

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.866	9.773	9.973	759.070	712.500	6.5
2,4-D	8.963	8.869	9.069	698.570	705.000	-0.9
2,4-DCAA	7.720	7.624	7.824	719.590	750.000	-4.1