

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM **Case No.:** P5117 **SAS No.:** P5117 **SDG NO.:** P5117

Continuing Calib Date: 12/06/2024 **Initial Calibration Date(s):** 11/27/2024 11/27/2024

Continuing Calib Time: 18:11 **Initial Calibration Time(s):** 11:29 12:24

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.09	9.09	8.99	9.19	0.00
Tetrachloro-m-xylene	3.56	3.56	3.46	3.66	0.00
gamma-BHC (Lindane)	4.34	4.34	4.24	4.44	0.00
Heptachlor	4.94	4.94	4.84	5.04	0.00
Heptachlor epoxide	5.70	5.70	5.60	5.80	0.00
Endrin	6.58	6.58	6.48	6.68	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5117 SAS No.: P5117 SDG NO.: P5117

Continuing Calib Date: 12/06/2024 Initial Calibration Date(s): 11/27/2024 11/27/2024

Continuing Calib Time: 18:11 Initial Calibration Time(s): 11:29 12:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.09	8.09	7.99	8.19	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
gamma-BHC (Lindane)	3.73	3.74	3.64	3.84	0.01
Heptachlor	4.09	4.09	3.99	4.19	0.00
Heptachlor epoxide	4.88	4.88	4.78	4.98	0.00
Endrin	5.80	5.80	5.70	5.90	0.00
Methoxychlor	6.77	6.77	6.67	6.87	0.00