

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5117 SAS No.: P5117 SDG NO.: P5117

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/27/2024 11/27/2024

Client Sample No.: CCAL01 Date Analyzed: 12/06/2024

Lab Sample No.: PSTDCCC050 Data File : PD087057.D Time Analyzed: 14:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.087	8.989	9.189	45.880	50.000	-8.2
Endrin	6.584	6.484	6.684	49.160	50.000	-1.7
gamma-BHC (Lindane)	4.338	4.238	4.438	53.280	50.000	6.6
Heptachlor	4.938	4.838	5.038	51.910	50.000	3.8
Heptachlor epoxide	5.701	5.600	5.800	51.380	50.000	2.8
Methoxychlor	7.504	7.403	7.603	44.990	50.000	-10.0
Tetrachloro-m-xylene	3.557	3.457	3.657	51.640	50.000	3.3

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5117 SAS No.: P5117 SDG NO.: P5117

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/27/2024 11/27/2024

Client Sample No.: CCAL01 Date Analyzed: 12/06/2024

Lab Sample No.: PSTDCCC050 Data File : PD087057.D Time Analyzed: 14:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.087	7.987	8.187	45.650	50.000	-8.7
Endrin	5.798	5.699	5.899	48.840	50.000	-2.3
gamma-BHC (Lindane)	3.734	3.635	3.835	51.400	50.000	2.8
Heptachlor	4.090	3.990	4.190	49.990	50.000	0.0
Heptachlor epoxide	4.881	4.781	4.981	50.080	50.000	0.2
Methoxychlor	6.765	6.667	6.867	46.620	50.000	-6.8
Tetrachloro-m-xylene	2.884	2.784	2.984	51.250	50.000	2.5