

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5117 SAS No.: P5117 SDG NO.: P5117

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/27/2024 11/27/2024

Client Sample No.: CCAL02 Date Analyzed: 12/06/2024

Lab Sample No.: PSTDCCC050 Data File : PD087071.D Time Analyzed: 18:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.088	8.989	9.189	46.760	50.000	-6.5
Endrin	6.584	6.484	6.684	49.110	50.000	-1.8
gamma-BHC (Lindane)	4.338	4.238	4.438	53.040	50.000	6.1
Heptachlor	4.939	4.838	5.038	51.360	50.000	2.7
Heptachlor epoxide	5.700	5.600	5.800	51.220	50.000	2.4
Methoxychlor	7.503	7.403	7.603	44.410	50.000	-11.2
Tetrachloro-m-xylene	3.557	3.457	3.657	51.600	50.000	3.2

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5117 SAS No.: P5117 SDG NO.: P5117

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/27/2024 11/27/2024

Client Sample No.: CCAL02 Date Analyzed: 12/06/2024

Lab Sample No.: PSTDCCC050 Data File : PD087071.D Time Analyzed: 18:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.086	7.987	8.187	46.390	50.000	-7.2
Endrin	5.798	5.699	5.899	48.400	50.000	-3.2
gamma-BHC (Lindane)	3.734	3.635	3.835	51.160	50.000	2.3
Heptachlor	4.089	3.990	4.190	49.530	50.000	-0.9
Heptachlor epoxide	4.880	4.781	4.981	49.660	50.000	-0.7
Methoxychlor	6.766	6.667	6.867	45.880	50.000	-8.2
Tetrachloro-m-xylene	2.883	2.784	2.984	50.830	50.000	1.7