

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120924\
 Data File : BF140793.D
 Acq On : 09 Dec 2024 16:54
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Dec 09 18:36:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.863	152	111338	20.000	ng	-0.01
21) Naphthalene-d8	8.145	136	407852	20.000	ng	-0.01
39) Acenaphthene-d10	9.904	164	228841	20.000	ng	-0.01
64) Phenanthrene-d10	11.398	188	436884	20.000	ng	0.00
76) Chrysene-d12	14.051	240	251461	20.000	ng	0.00
86) Perylene-d12	15.557	264	241288	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	546695	83.779	ng	0.00
7) Phenol-d6	6.516	99	715376	82.925	ng	0.00
23) Nitrobenzene-d5	7.434	82	666058	83.533	ng	0.00
42) 2,4,6-Tribromophenol	10.698	330	211482	86.408	ng	0.00
45) 2-Fluorobiphenyl	9.222	172	1271155	82.763	ng	-0.01
79) Terphenyl-d14	12.980	244	1340625	83.016	ng	-0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.628	88	113183	41.253	ng	97
3) Pyridine	3.404	79	261113	43.255	ng	99
4) n-Nitrosodimethylamine	3.369	42	140273	39.537	ng	91
6) Aniline	6.534	93	247715	40.796	ng	# 71
8) 2-Chlorophenol	6.657	128	296851	42.284	ng	99
9) Benzaldehyde	6.422	77	179459	39.296	ng	99
10) Phenol	6.528	94	366410	41.590	ng	98
11) bis(2-Chloroethyl)ether	6.604	93	285071	42.284	ng	98
12) 1,3-Dichlorobenzene	6.804	146	331247	41.991	ng	99
13) 1,4-Dichlorobenzene	6.881	146	328765	41.161	ng	99
14) 1,2-Dichlorobenzene	7.040	146	313249	41.853	ng	99
15) Benzyl Alcohol	7.016	79	250728	39.191	ng	97
16) 2,2'-oxybis(1-Chloropr...	7.140	45	319710	40.141	ng	71
17) 2-Methylphenol	7.134	107	232545	41.380	ng	95
18) Hexachloroethane	7.375	117	122849	41.181	ng	100
19) n-Nitroso-di-n-propyla...	7.287	70	202526	39.723	ng	96
20) 3+4-Methylphenols	7.287	107	299394	41.448	ng	# 85
22) Acetophenone	7.281	105	409520	41.186	ng	97
24) Nitrobenzene	7.457	77	340292	41.293	ng	97
25) Isophorone	7.692	82	555956	41.803	ng	99
26) 2-Nitrophenol	7.769	139	160033	43.820	ng	96
27) 2,4-Dimethylphenol	7.810	122	180002	41.194	ng	98
28) bis(2-Chloroethoxy)met...	7.892	93	344075	42.468	ng	99
29) 2,4-Dichlorophenol	8.022	162	250923	43.337	ng	98
30) 1,2,4-Trichlorobenzene	8.087	180	278158	42.076	ng	100
31) Naphthalene	8.169	128	890300	42.379	ng	99
32) Benzoic acid	7.969	122	179000	47.213	ng	98
33) 4-Chloroaniline	8.228	127	258600	41.114	ng	99
34) Hexachlorobutadiene	8.281	225	182142	41.597	ng	98
35) Caprolactam	8.604	113	58787	32.808	ng	98
36) 4-Chloro-3-methylphenol	8.716	107	274503	42.344	ng	99
37) 2-Methylnaphthalene	8.857	142	565502	42.381	ng	98
38) 1-Methylnaphthalene	8.957	142	554906	42.427	ng	100
40) 1,2,4,5-Tetrachloroben...	9.028	216	283106	42.259	ng	99
41) Hexachlorocyclopentadiene	9.010	237	51965	38.655	ng	99
43) 2,4,6-Trichlorophenol	9.145	196	179504	42.701	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120924\
 Data File : BF140793.D
 Acq On : 09 Dec 2024 16:54
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Quant Time: Dec 09 18:36:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	191015	41.868	ng	97
46) 1,1'-Biphenyl	9.322	154	712274	41.689	ng	99
47) 2-Chloronaphthalene	9.351	162	548428	42.363	ng	99
48) 2-Nitroaniline	9.457	65	173500	41.753	ng	95
49) Acenaphthylene	9.769	152	802935	41.049	ng	99
50) Dimethylphthalate	9.628	163	639675	42.443	ng	100
51) 2,6-Dinitrotoluene	9.698	165	145468	42.558	ng	94
52) Acenaphthene	9.939	154	516945	41.593	ng	99
53) 3-Nitroaniline	9.869	138	144912	43.347	ng	98
54) 2,4-Dinitrophenol	9.986	184	88646	53.002	ng	95
55) Dibenzofuran	10.110	168	783038	41.305	ng	99
56) 4-Nitrophenol	10.051	139	123343	54.099	ng	95
57) 2,4-Dinitrotoluene	10.104	165	197930	43.570	ng	96
58) Fluorene	10.457	166	644283	42.340	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.239	232	162969	46.479	ng	91
60) Diethylphthalate	10.322	149	651981	42.578	ng	100
61) 4-Chlorophenyl-phenyle...	10.445	204	322425	43.176	ng	96
62) 4-Nitroaniline	10.492	138	154265	43.997	ng	94
63) Azobenzene	10.604	77	605475	41.754	ng	98
65) 4,6-Dinitro-2-methylph...	10.522	198	109749	49.160	ng	93
66) n-Nitrosodiphenylamine	10.569	169	534148	41.343	ng	99
67) 4-Bromophenyl-phenylether	10.933	248	191210	42.498	ng	98
68) Hexachlorobenzene	11.004	284	221124	42.445	ng	97
69) Atrazine	11.092	200	124649	34.295	ng	99
70) Pentachlorophenol	11.210	266	87095	37.985	ng	99
71) Phenanthrene	11.422	178	871003	41.475	ng	99
72) Anthracene	11.475	178	858438	41.788	ng	100
73) Carbazole	11.633	167	829167	41.957	ng	100
74) Di-n-butylphthalate	11.951	149	980776	42.988	ng	99
75) Fluoranthene	12.616	202	963587	42.260	ng	100
77) Benzidine	12.739	184	220453	30.153	ng	99
78) Pyrene	12.845	202	974808	41.926	ng	99
80) Butylbenzylphthalate	13.451	149	341631	40.788	ng	98
81) Benzo(a)anthracene	14.039	228	722318	43.394	ng	100
82) 3,3'-Dichlorobenzidine	13.998	252	219006	43.822	ng	100
83) Chrysene	14.080	228	612266	40.226	ng	99
84) Bis(2-ethylhexyl)phtha...	14.010	149	415362	39.273	ng	99
85) Di-n-octyl phthalate	14.639	149	626554	43.349	ng	98
87) Indeno(1,2,3-cd)pyrene	17.086	276	748155	47.579	ng	97
88) Benzo(b)fluoranthene	15.115	252	637387	42.056	ng	99
89) Benzo(k)fluoranthene	15.145	252	529858	39.952	ng	99
90) Benzo(a)pyrene	15.492	252	516945	41.950	ng	98
91) Dibenzo(a,h)anthracene	17.092	278	618905	48.060	ng	99
92) Benzo(g,h,i)perylene	17.539	276	641395	48.809	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 09 18:36:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 21 15:23:48 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.863 min Scan# 81
 Delta R.T. -0.012 min
 Lab File: BF140793.D
 Acq: 09 Dec 2024 16:54




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:152 Resp: 111338
 Ion Ratio Lower Upper
 152 100
 150 158.9 128.5 192.7
 115 61.9 50.2 75.4



#2
 1,4-Dioxane
 Concen: 41.253 ng
 RT: 2.628 min Scan# 91
 Delta R.T. -0.023 min
 Lab File: BF140793.D
 Acq: 09 Dec 2024 16:54

Tgt Ion: 88 Resp: 113183
 Ion Ratio Lower Upper
 88 100
 58 68.3 57.4 86.2
 43 30.1 24.3 36.5





#3

Pyridine

Concen: 43.255 ng

RT: 3.404 min Scan# 21

Delta R.T. -0.018 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion: 79 Resp: 261113

Ion Ratio Lower Upper

79 100

52 65.4 52.1 78.1

51 34.1 27.8 41.8



#4

n-Nitrosodimethylamine

Concen: 39.537 ng

RT: 3.369 min Scan# 217

Delta R.T. -0.012 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54




Tgt Ion: 42 Resp: 140273


Ion Ratio Lower Upper

42 100

74 127.0 93.8 140.6

44 6.4 4.6 6.8









#5
2-Fluorophenol
Concen: 83.779 ng
RT: 5.493 min Scan# 51
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:112 Resp: 546695
Ion Ratio Lower Upper
112 100
64 59.1 49.2 73.8
63 31.9 27.0 40.4



#6
Aniline
Concen: 40.796 ng
RT: 6.534 min Scan# 755
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion: 93 Resp: 247715
Ion Ratio Lower Upper
93 100
66 95.3 56.5 84.7#
65 57.7 32.9 49.3#





#7

Phenol-d6

Concen: 82.925 ng

RT: 6.516 min Scan# 715

Delta R.T. 0.000 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion: 99 Resp: 715376

Ion Ratio Lower Upper

99 100

42 17.6 15.4 23.0

71 36.1 30.6 46.0



#8

2-Chlorophenol

Concen: 42.284 ng

RT: 6.657 min Scan# 776

Delta R.T. -0.006 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54




Tgt Ion:128 Resp: 296851


Ion Ratio Lower Upper

128 100

130 32.3 12.3 52.3

64 48.6 30.2 70.2









#9
Benzaldehyde
Concen: 39.296 ng
RT: 6.422 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 77 Resp: 179459
Ion Ratio Lower Upper
77 100
105 97.0 75.6 115.6
106 92.0 72.0 112.0



#10
Phenol
Concen: 41.590 ng
RT: 6.528 min Scan# 754
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion: 94 Resp: 366410
Ion Ratio Lower Upper
94 100
65 37.4 17.2 57.2
66 61.8 39.2 79.2









#11
bis(2-Chloroethyl)ether
Concen: 42.284 ng
RT: 6.604 min Scan# 768
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 93 Resp: 285071
Ion Ratio Lower Upper
93 100
63 71.0 52.8 92.8
95 32.1 12.3 52.3



#12
1,3-Dichlorobenzene
Concen: 41.991 ng
RT: 6.804 min Scan# 801
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:146 Resp: 331247
Ion Ratio Lower Upper
146 100
148 63.7 51.3 76.9
75 33.6 28.2 42.2









#13
1,4-Dichlorobenzene
Concen: 41.161 ng
RT: 6.881 min Scan# 816
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:146 Resp: 328765
Ion Ratio Lower Upper
146 100
148 63.8 51.0 76.6
111 44.9 35.3 52.9



#14
1,2-Dichlorobenzene
Concen: 41.853 ng
RT: 7.040 min Scan# 841
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:146 Resp: 313249
Ion Ratio Lower Upper
146 100
148 63.8 50.5 75.7
111 44.8 36.7 55.1









#15
Benzyl Alcohol
Concen: 39.191 ng
RT: 7.016 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 79 Resp: 250728
Ion Ratio Lower Upper
79 100
108 72.3 55.0 82.6
77 63.2 50.1 75.1



#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.141 ng
RT: 7.140 min Scan# 858
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion: 45 Resp: 319710
Ion Ratio Lower Upper
45 100
77 46.0 10.7 50.7
79 42.3 7.0 47.0









#17
2-Methylphenol
Concen: 41.380 ng
RT: 7.134 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	107	Resp:	232545
Ion	Ratio	Lower	Upper
107	100		
108	113.2	91.7	137.5
77	69.2	49.3	73.9
79	64.7	47.1	70.7



#18
Hexachloroethane
Concen: 41.181 ng
RT: 7.375 min Scan# 898
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:	117	Resp:	122849
Ion	Ratio	Lower	Upper
117	100		
119	96.8	77.1	115.7
201	101.2	81.1	121.7









#19
n-Nitroso-di-n-propylamine
Concen: 39.723 ng
RT: 7.287 min Scan# 883
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion: 70 Resp: 202526




Ion	Ratio	Lower	Upper
70	100		
42	55.7	47.6	71.4
101	10.4	7.6	11.4
130	21.5	16.2	24.4




#20
3+4-Methylphenols
Concen: 41.448 ng
RT: 7.287 min Scan# 883
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion: 107 Resp: 299394

Ion	Ratio	Lower	Upper
107	100		
108	90.4	69.9	109.9
77	104.1	117.7	157.7#
79	32.2	13.4	53.4









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.145 min Scan# 1031
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:136 Resp: 407852




Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.6	13.0
54	7.1	5.8	8.8
68	5.5	4.1	6.1




#22
Acetophenone
Concen: 41.186 ng
RT: 7.281 min Scan# 882
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:105 Resp: 409520

Ion	Ratio	Lower	Upper
105	100		
71	7.8	6.2	9.2
51	33.8	29.0	43.6
120	22.0	17.4	26.0









#23
Nitrobenzene-d5
Concen: 83.533 ng
RT: 7.434 min Scan# 909
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 82 Resp: 666058
Ion Ratio Lower Upper
82 100
128 42.7 33.0 49.4
54 47.6 39.5 59.3



#24
Nitrobenzene
Concen: 41.293 ng
RT: 7.457 min Scan# 912
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion: 77 Resp: 340292
Ion Ratio Lower Upper
77 100
123 45.7 34.3 51.5
65 13.7 11.2 16.8









#25
Isophorone
Concen: 41.803 ng
RT: 7.692 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 82 Resp: 555956
Ion Ratio Lower Upper
82 100
95 7.9 6.4 9.6
138 17.6 13.8 20.6



#26
2-Nitrophenol
Concen: 43.820 ng
RT: 7.769 min Scan# 965
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:139 Resp: 160033
Ion Ratio Lower Upper
139 100
109 33.3 28.0 42.0
65 52.6 44.6 67.0









#27
2,4-Dimethylphenol
Concen: 41.194 ng
RT: 7.810 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:122 Resp: 180002
Ion Ratio Lower Upper
122 100
107 118.2 96.8 145.2
121 59.1 47.4 71.0



#28
bis(2-Chloroethoxy)methane
Concen: 42.468 ng
RT: 7.892 min Scan# 986
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion: 93 Resp: 344075
Ion Ratio Lower Upper
93 100
95 32.3 26.1 39.1
123 10.7 9.1 13.7





#29

2,4-Dichlorophenol

Concen: 43.337 ng

RT: 8.022 min Scan# 1008

Delta R.T. 0.006 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:162 Resp: 250923

Ion Ratio Lower Upper

162 100

164 63.6 44.8 84.8

98 38.9 21.2 61.2



#30

1,2,4-Trichlorobenzene

Concen: 42.076 ng

RT: 8.087 min Scan# 1019

Delta R.T. -0.012 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54




Tgt Ion:180 Resp: 278158


Ion Ratio Lower Upper

180 100

182 95.5 76.4 114.6

145 32.5 25.9 38.9









#31
Naphthalene
Concen: 42.379 ng
RT: 8.169 min Scan# 1033
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:128 Resp: 890300
Ion Ratio Lower Upper
128 100
129 11.1 8.6 13.0
127 13.3 10.8 16.2



#32
Benzoic acid
Concen: 47.213 ng
RT: 7.969 min Scan# 999
Delta R.T. 0.030 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:122 Resp: 179000
Ion Ratio Lower Upper
122 100
105 131.7 109.3 149.3
77 101.4 82.3 122.3









#33
4-Chloroaniline
Concen: 41.114 ng
RT: 8.228 min Scan# 1043
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	127	Resp:	258600
Ion	Ratio	Lower	Upper
127	100		
129	32.2	26.0	39.0
65	31.2	26.0	39.0
92	20.2	16.7	25.1



#34
Hexachlorobutadiene
Concen: 41.597 ng
RT: 8.281 min Scan# 1052
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:	225	Resp:	182142
Ion	Ratio	Lower	Upper
225	100		
223	64.2	49.8	74.8
227	64.6	51.3	76.9









#35
 Caprolactam
 Concen: 32.808 ng
 RT: 8.604 min Scan# 1107
 Delta R.T. 0.000 min
 Lab File: BF140793.D
 Acq: 09 Dec 2024 16:54




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:113 Resp: 58787
 Ion Ratio Lower Upper
 113 100
 55 156.9 137.7 177.7
 56 121.8 106.4 146.4



#36
 4-Chloro-3-methylphenol
 Concen: 42.344 ng
 RT: 8.716 min Scan# 1126
 Delta R.T. 0.000 min
 Lab File: BF140793.D
 Acq: 09 Dec 2024 16:54

Tgt Ion:107 Resp: 274503
 Ion Ratio Lower Upper
 107 100
 144 24.3 18.7 28.1
 142 73.7 58.3 87.5





#37

2-Methylnaphthalene

Concen: 42.381 ng

RT: 8.857 min Scan# 1152

Delta R.T. -0.012 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040EC

Tgt Ion:142 Resp: 565502

Ion Ratio Lower Upper

142 100

141 87.4 71.0 106.6



#38

1-Methylnaphthalene

Concen: 42.427 ng

RT: 8.957 min Scan# 1167

Delta R.T. -0.012 min

Lab File: BF140793.D




Acq: 09 Dec 2024 16:54


Tgt Ion:142 Resp: 554906

Ion Ratio Lower Upper

142 100

141 91.4 73.3 109.9









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.904 min Scan# 1179
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:164 Resp: 228841
Ion Ratio Lower Upper
164 100
162 100.5 80.6 120.8
160 45.2 36.2 54.4



#40
1,2,4,5-Tetrachlorobenzene
Concen: 42.259 ng
RT: 9.028 min Scan# 1179
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:216 Resp: 283106
Ion Ratio Lower Upper
216 100
214 78.7 63.0 94.4
179 20.9 17.5 26.3
108 17.0 14.7 22.1









#41
Hexachlorocyclopentadiene
Concen: 38.655 ng
RT: 9.010 min Scan# 1177
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:237 Resp: 51965
Ion Ratio Lower Upper
237 100
235 61.7 41.1 81.1
272 12.8 0.0 32.6



#42
2,4,6-Tribromophenol
Concen: 86.408 ng
RT: 10.698 min Scan# 1463
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:330 Resp: 211482
Ion Ratio Lower Upper
330 100
332 97.5 76.9 115.3
141 31.7 26.7 40.1





#43

2,4,6-Trichlorophenol

Concen: 42.701 ng

RT: 9.145 min Scan# 1199

Delta R.T. -0.006 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:196 Resp: 179504

Ion Ratio Lower Upper

196 100

198 95.3 76.8 115.2

200 30.5 24.3 36.5



#44

2,4,5-Trichlorophenol

Concen: 41.868 ng

RT: 9.198 min Scan# 1208

Delta R.T. 0.006 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Tgt Ion:196 Resp: 191015




Ion Ratio Lower Upper


196 100

198 96.5 77.2 115.8

97 53.3 47.1 70.7

132 29.0 24.1 36.1









#45
2-Fluorobiphenyl
Concen: 82.763 ng
RT: 9.222 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:172 Resp: 1271155
Ion Ratio Lower Upper
172 100
171 35.5 28.4 42.6
170 23.5 19.0 28.6



#46
1,1'-Biphenyl
Concen: 41.689 ng
RT: 9.322 min Scan# 1229
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:154 Resp: 712274
Ion Ratio Lower Upper
154 100
153 40.8 20.4 60.4
76 11.9 0.0 31.2





#47

2-Chloronaphthalene

Concen: 42.363 ng

RT: 9.351 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:162 Resp: 548428

Ion Ratio Lower Upper

162 100

127 43.2 35.7 53.5

164 32.8 26.1 39.1



#48

2-Nitroaniline

Concen: 41.753 ng

RT: 9.457 min Scan# 1252

Delta R.T. 0.000 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54




Tgt Ion: 65 Resp: 173500


Ion Ratio Lower Upper

65 100

92 71.9 56.1 84.1

138 109.0 81.0 121.6





#49

Acenaphthylene

Concen: 41.049 ng

RT: 9.769 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:152 Resp: 802935

Ion Ratio Lower Upper

152 100

151 20.3 15.9 23.9

153 13.2 10.6 16.0



#50

Dimethylphthalate

Concen: 42.443 ng

RT: 9.628 min Scan# 1281

Delta R.T. -0.006 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54




Tgt Ion:163 Resp: 639675


Ion Ratio Lower Upper

163 100

194 3.9 3.1 4.7

164 10.0 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 42.558 ng
RT: 9.698 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:165 Resp: 145468
Ion Ratio Lower Upper
165 100
63 57.7 49.9 74.9
89 58.3 50.9 76.3



#52
Acenaphthene
Concen: 41.593 ng
RT: 9.939 min Scan# 1334
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:154 Resp: 516945
Ion Ratio Lower Upper
154 100
153 113.0 89.7 134.5
152 53.3 42.5 63.7









#53
3-Nitroaniline
Concen: 43.347 ng
RT: 9.869 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:138 Resp: 144912
Ion Ratio Lower Upper
138 100
108 11.3 9.2 13.8
92 123.0 100.6 150.8



#54
2,4-Dinitrophenol
Concen: 53.002 ng
RT: 9.986 min Scan# 1342
Delta R.T. 0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:184 Resp: 88646
Ion Ratio Lower Upper
184 100
63 62.8 54.1 81.1
154 68.3 57.4 86.0









#55
Dibenzenofuran
Concen: 41.305 ng
RT: 10.110 min Scan# 1363
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:168 Resp: 783038
Ion Ratio Lower Upper
168 100
139 39.7 31.3 46.9
169 13.6 10.5 15.7



#56
4-Nitrophenol
Concen: 54.099 ng
RT: 10.051 min Scan# 1353
Delta R.T. 0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:139 Resp: 123343
Ion Ratio Lower Upper
139 100
109 82.9 68.4 108.4
65 104.9 88.6 128.6









#57
2,4-Dinitrotoluene
Concen: 43.570 ng
RT: 10.104 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:165 Resp: 197930
Ion Ratio Lower Upper
165 100
63 58.3 41.7 62.5
89 82.8 65.8 98.8
182 2.8 2.1 3.1



#58
Fluorene
Concen: 42.340 ng
RT: 10.457 min Scan# 1422
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:166 Resp: 644283
Ion Ratio Lower Upper
166 100
165 93.6 74.6 111.8
167 13.3 10.6 16.0









#59
2,3,4,6-Tetrachlorophenol
Concen: 46.479 ng
RT: 10.239 min Scan# 1385
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:232 Resp: 162969
Ion Ratio Lower Upper
232 100
131 42.7 39.6 59.4
130 2.3 1.9 2.9
166 26.9 25.3 37.9



#60
Diethylphthalate
Concen: 42.578 ng
RT: 10.322 min Scan# 1399
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:149 Resp: 651981
Ion Ratio Lower Upper
149 100
177 21.7 17.1 25.7
150 12.6 10.1 15.1









#61
4-Chlorophenyl-phenylether
Concen: 43.176 ng
RT: 10.445 min Scan# 1421
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:204 Resp: 322425
Ion Ratio Lower Upper
204 100
206 32.2 26.6 40.0
141 59.7 51.1 76.7



#62
4-Nitroaniline
Concen: 43.997 ng
RT: 10.492 min Scan# 1428
Delta R.T. 0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:138 Resp: 154265
Ion Ratio Lower Upper
138 100
92 52.7 34.5 74.5
108 82.4 70.6 110.6





#63

Azobenzene

Concen: 41.754 ng

RT: 10.604 min Scan# 1447

Delta R.T. -0.012 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC

Tgt Ion: 77 Resp: 605475





Ion Ratio Lower Upper

77 100

182 24.8 5.3 45.3

105 20.9 0.3 40.3

51 28.2 9.4 49.4



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.398 min Scan# 1582

Delta R.T. -0.006 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54




Tgt Ion: 188 Resp: 436884


Ion Ratio Lower Upper

188 100

94 8.2 6.4 9.6

80 8.6 6.9 10.3









#65
4,6-Dinitro-2-methylphenol
Concen: 49.160 ng
RT: 10.522 min Scan# 1433
Delta R.T. 0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:198 Resp: 109749
Ion Ratio Lower Upper
198 100
51 42.8 29.4 69.4
105 43.1 25.4 65.4



#66
n-Nitrosodiphenylamine
Concen: 41.343 ng
RT: 10.569 min Scan# 1441
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:169 Resp: 534148
Ion Ratio Lower Upper
169 100
168 67.0 53.2 79.8
167 36.9 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 42.498 ng
RT: 10.933 min Scan# 1503
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:248 Resp: 191210
Ion Ratio Lower Upper
248 100
250 97.6 79.8 119.8
141 79.5 64.7 97.1



#68
Hexachlorobenzene
Concen: 42.445 ng
RT: 11.004 min Scan# 1515
Delta R.T. -0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:284 Resp: 221124
Ion Ratio Lower Upper
284 100
142 34.4 25.3 37.9
249 31.1 24.8 37.2





#69

Atrazine

Concen: 34.295 ng

RT: 11.092 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:200 Resp: 124649

Ion Ratio Lower Upper

200 100

173 26.0 7.0 47.0

215 46.6 26.7 66.7



#70

Pentachlorophenol

Concen: 37.985 ng

RT: 11.210 min Scan# 1550

Delta R.T. -0.000 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54




Tgt Ion:266 Resp: 87095


Ion Ratio Lower Upper

266 100

268 65.8 53.7 80.5

264 62.3 49.8 74.8









#71
Phenanthrene
Concen: 41.475 ng
RT: 11.422 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:178 Resp: 871003
Ion Ratio Lower Upper
178 100
176 19.3 15.8 23.6
179 15.5 12.4 18.6



#72
Anthracene
Concen: 41.788 ng
RT: 11.475 min Scan# 1595
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:178 Resp: 858438
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.4
179 15.3 12.2 18.4









#73
 Carbazole
 Concen: 41.957 ng
 RT: 11.633 min Scan# 1622
 Delta R.T. -0.006 min
 Lab File: BF140793.D
 Acq: 09 Dec 2024 16:54




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:167 Resp: 829167
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.2 25.8
 139 13.1 10.6 15.8



#74
 Di-n-butylphthalate
 Concen: 42.988 ng
 RT: 11.951 min Scan# 1676
 Delta R.T. -0.006 min
 Lab File: BF140793.D
 Acq: 09 Dec 2024 16:54

Tgt Ion:149 Resp: 980776
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.5 11.3
 104 4.9 4.1 6.1





#75

Fluoranthene

Concen: 42.260 ng

RT: 12.616 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:202 Resp: 963587

Ion Ratio Lower Upper

202 100

101 9.3 0.0 28.9

203 17.7 0.0 37.6



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.051 min Scan# 2033

Delta R.T. 0.000 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54




Tgt Ion:240 Resp: 251461


Ion Ratio Lower Upper

240 100

120 9.5 7.3 10.9

236 25.9 20.6 31.0









#77
Benzidine
Concen: 30.153 ng
RT: 12.739 min Scan# 1810
Delta R.T. 0.000 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:184 Resp: 220453
Ion Ratio Lower Upper
184 100
185 14.1 11.7 17.5
183 10.7 8.8 13.2



#78
Pyrene
Concen: 41.926 ng
RT: 12.845 min Scan# 1828
Delta R.T. -0.006 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:202 Resp: 974808
Ion Ratio Lower Upper
202 100
200 20.0 16.5 24.7
203 17.3 14.2 21.4





#79

Terphenyl-d14

Concen: 83.016 ng

RT: 12.980 min Scan# 1851

Delta R.T. -0.012 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:244 Resp: 1340625

Ion Ratio Lower Upper

244 100

212 7.1 5.8 8.8

122 10.5 8.0 12.0



#80

Butylbenzylphthalate

Concen: 40.788 ng

RT: 13.451 min Scan# 1931

Delta R.T. -0.012 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54




Tgt Ion:149 Resp: 341631


Ion Ratio Lower Upper

149 100

91 72.7 56.2 84.2

206 20.7 16.4 24.6





#81

Benzo(a)anthracene

Concen: 43.394 ng

RT: 14.039 min Scan# 2031

Delta R.T. 0.000 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:228 Resp: 722318

Ion Ratio Lower Upper

228 100

226 27.1 21.8 32.8

229 19.7 15.8 23.6



#82

3,3'-Dichlorobenzidine

Concen: 43.822 ng

RT: 13.998 min Scan# 2024

Delta R.T. 0.000 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54




Tgt Ion:252 Resp: 219006


Ion Ratio Lower Upper

252 100

254 63.8 50.8 76.2

126 10.6 8.6 13.0





#83

Chrysene

Concen: 40.226 ng

RT: 14.080 min Scan# 2038

Delta R.T. 0.000 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:228 Resp: 612266

Ion Ratio Lower Upper

228 100

226 29.4 24.4 36.6

229 19.5 15.8 23.6



#84

Bis(2-ethylhexyl)phthalate

Concen: 39.273 ng

RT: 14.010 min Scan# 2026

Delta R.T. -0.006 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54




Tgt Ion:149 Resp: 415362


Ion Ratio Lower Upper

149 100

167 26.9 21.2 31.8

279 4.0 3.2 4.8





#85

Di-n-octyl phthalate

Concen: 43.349 ng

RT: 14.639 min Scan# 2132

Delta R.T. 0.006 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:149 Resp: 626554

Ion Ratio Lower Upper

149 100

167 1.7 1.3 1.9

43 8.8 7.7 11.5



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.557 min Scan# 2289

Delta R.T. 0.018 min

Lab File: BF140793.D

Acq: 09 Dec 2024 16:54




Tgt Ion:264 Resp: 241288


Ion Ratio Lower Upper

264 100

260 23.4 19.0 28.6

265 21.9 16.6 25.0









#87
Indeno(1,2,3-cd)pyrene
Concen: 47.579 ng
RT: 17.086 min Scan# 2114
Delta R.T. 0.035 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:276 Resp: 748155
Ion Ratio Lower Upper
276 100
138 21.6 15.5 23.3
277 25.4 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 42.056 ng
RT: 15.115 min Scan# 2214
Delta R.T. 0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:252 Resp: 637387
Ion Ratio Lower Upper
252 100
253 21.9 17.2 25.8
125 8.9 6.7 10.1









#89
Benzo(k)fluoranthene
Concen: 39.952 ng
RT: 15.145 min Scan# 2219
Delta R.T. 0.012 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:252 Resp: 529858
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 9.0 6.5 9.7



#90
Benzo(a)pyrene
Concen: 41.950 ng
RT: 15.492 min Scan# 2278
Delta R.T. 0.018 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:252 Resp: 516945
Ion Ratio Lower Upper
252 100
253 22.2 17.3 25.9
125 9.9 7.0 10.6









#91
Dibenzo(a,h)anthracene
Concen: 48.060 ng
RT: 17.092 min Scan# 21
Delta R.T. 0.029 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:278 Resp: 618905
Ion Ratio Lower Upper
278 100
139 13.6 10.2 15.4
279 23.2 18.9 28.3



#92
Benzo(g,h,i)perylene
Concen: 48.809 ng
RT: 17.539 min Scan# 2626
Delta R.T. 0.029 min
Lab File: BF140793.D
Acq: 09 Dec 2024 16:54

Tgt Ion:276 Resp: 641395
Ion Ratio Lower Upper
276 100
277 23.8 18.9 28.3
138 19.1 13.2 19.8

