

Surrogate Summary

SW-846

SDG No.: P5117

Client: Weston Solutions

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P5095-04MS	MH-764MS	2-Fluorophenol	150	128	86		19	119
		Phenol-d6	150	118	79		10	130
		Nitrobenzene-d5	100	88.2	88		44	120
		2-Fluorobiphenyl	100	89.2	89		44	119
		2,4,6-Tribromophenol	150	134	89		43	140
		Terphenyl-d14	100	100	100		50	134
P5095-04MSD	MH-764MSD	2-Fluorophenol	150	150	100		19	119
		Phenol-d6	150	136	91		10	130
		Nitrobenzene-d5	100	108	108		44	120
		2-Fluorobiphenyl	100	112	112		44	119
		2,4,6-Tribromophenol	150	166	111		43	140
		Terphenyl-d14	100	117	117		50	134
P5117-02	TAPIAL2-IDW-SOIL-120424-00-T2	2-Fluorophenol	150	127	85		19	119
		Phenol-d6	150	117	78		10	130
		Nitrobenzene-d5	100	89.3	89		44	120
		2-Fluorobiphenyl	100	93.9	94		44	119
		2,4,6-Tribromophenol	150	138	92		43	140
		Terphenyl-d14	100	93.1	93		50	134
PB165390TB	PB165390TB	2-Fluorophenol	150	149	100		19	119
		Phenol-d6	150	147	98		10	130
		Nitrobenzene-d5	100	98.3	98		44	120
		2-Fluorobiphenyl	100	101	101		44	119
		2,4,6-Tribromophenol	150	153	102		43	140
		Terphenyl-d14	100	90.7	91		50	134
PB165435BL	PB165435BL	2-Fluorophenol	150	159	106		19	119
		Phenol-d6	150	154	103		10	130
		Nitrobenzene-d5	100	103	103		44	120
		2-Fluorobiphenyl	100	107	107		44	119
		2,4,6-Tribromophenol	150	159	106		43	140
		Terphenyl-d14	100	99.0	99		50	134
PB165435BS	PB165435BS	2-Fluorophenol	150	141	94		19	119
		Phenol-d6	150	138	92		10	130
		Nitrobenzene-d5	100	94.0	94		44	120
		2-Fluorobiphenyl	100	97.2	97		44	119
		2,4,6-Tribromophenol	150	148	99		43	140
		Terphenyl-d14	100	94.4	94		50	134