

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5117

Client: Weston Solutions

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P5095-04MSD	Client Sample ID:	MH-764MSD					DataFile:	BF140792.D		
Pyridine	500	0	390	ug/L	78		20		10	153	20
1,4-Dichlorobenzene	500	0	460	ug/L	92		11		29	112	20
2-Methylphenol	500	0	530	ug/L	106		14		30	117	20
3+4-Methylphenols	500	0	540	ug/L	108		12		29	110	20
Hexachloroethane	500	0	450	ug/L	90		14		21	115	20
Nitrobenzene	500	0	510	ug/L	102		22	*	45	121	20
Hexachlorobutadiene	500	0	480	ug/L	96		16		22	124	20
2,4,6-Trichlorophenol	500	0	550	ug/L	110		20		50	125	20
2,4,5-Trichlorophenol	500	0	540	ug/L	108		20		53	123	20
2,4-Dinitrotoluene	500	0	550	ug/L	110		18		57	128	20
Hexachlorobenzene	500	0	550	ug/L	110		18		53	125	20
Pentachlorophenol	1000	0	990	ug/L	99		12		35	138	20