

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WEST04

Lab Code: CHEM Case No.: P5117 SAS No.: P5117 SDG No.: P5117

Instrument ID: MSVOA_N Calibration Date/Time: 12/09/2024 10:32

Lab File ID: VN085142.D Init. Calib. Date(s): 11/22/2024 11/22/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:22 13:45

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.589	0.495		-15.96	20
1,1-Dichloroethene	0.531	0.486		-8.48	20
2-Butanone	0.339	0.351		3.54	20
Carbon Tetrachloride	0.527	0.559		6.07	20
Chloroform	1.146	1.087		-5.15	20
Benzene	1.445	1.457		0.9	20
1,2-Dichloroethane	0.494	0.498		0.81	20
Trichloroethene	0.336	0.347		3.27	20
Tetrachloroethene	0.337	0.355		5.34	20
Chlorobenzene	1.134	1.124	0.3	-0.88	20
1,2-Dichloroethane-d4	0.712	0.678		-4.78	20
Dibromofluoromethane	0.337	0.350		3.86	20
Toluene-d8	1.208	1.232		1.99	20
4-Bromofluorobenzene	0.452	0.465		2.88	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.